
**Passive personal neutron dosimeters —
Performance and test requirements**

*Dosimètres individuels passifs pour les neutrons — Exigences de
fonctionnement et d'essai*

STANDARDSISO.COM : Click to view the full PDF of ISO 21909:2005



PDF disclaimer

This PDF file may contain embedded typefaces. In accordance with Adobe's licensing policy, this file may be printed or viewed but shall not be edited unless the typefaces which are embedded are licensed to and installed on the computer performing the editing. In downloading this file, parties accept therein the responsibility of not infringing Adobe's licensing policy. The ISO Central Secretariat accepts no liability in this area.

Adobe is a trademark of Adobe Systems Incorporated.

Details of the software products used to create this PDF file can be found in the General Info relative to the file; the PDF-creation parameters were optimized for printing. Every care has been taken to ensure that the file is suitable for use by ISO member bodies. In the unlikely event that a problem relating to it is found, please inform the Central Secretariat at the address given below.

STANDARDSISO.COM : Click to view the full PDF of ISO 21909:2005

© ISO 2005

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

Page

Foreword.....	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
3.1 General terms	1
3.2 Quantities	2
3.3 Calibration and evaluation	3
3.4 Terms with respect to nuclear track emulsion dosimeters (NTED)	6
3.5 Terms with respect to solid state nuclear track dosimeters (SSNTD)	7
3.6 Terms with respect to thermoluminescence albedo dosimeters (TLAD)	7
3.7 Terms with respect to superheated emulsion dosimeters (SED)	8
3.8 Terms with respect to ion chamber dosimeters with direct ion storage (ICD)	9
4 Units	10
5 General test conditions	10
5.1 Test conditions	10
5.2 Reference radiation	10
5.3 Test requirements	10
6 Performance requirements	10
7 Test methods	15
8 Identification and accompanying documentation	16
8.1 Individual marking	16
8.2 Collective marking	16
8.3 Accompanying documentation	16
9 Certification	16
Annex A (normative) Reference and standard test conditions	17
Annex B (normative) Confidence limits	18
Annex C (normative) Performance tests	21
Annex D (informative) Irradiation conditions	38
Annex E (informative) Performance of neutron dosimeters	39
Annex F (normative) Conversion tables	48
Bibliography	49

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 21909 was prepared by Technical Committee ISO/TC 85, *Nuclear energy*, Subcommittee SC 2, *Radiation protection*.

STANDARDSISO.COM : Click to view the full PDF of ISO 21909:2005

Introduction

ISO 21909 contains performance and test requirements for dosimeters to be used for the determination of personal dose equivalent, $H_p(10)$, in neutron fields with energies ranging from thermal to 20 MeV. In this energy range, reference neutron radiation are reported in ISO 8529-3. Reference radiation are required for the proper calibration of the dosimeters. This International Standard covers the following five classes of passive neutron detectors that can be used as a personal dosimeter in part or all of the above-mentioned neutron energy range:

- nuclear track emulsion dosimeters (NTED);
- solid state nuclear track dosimeters (SSNTD);
- thermoluminescence albedo dosimeters (TLAD);
- superheated emulsion dosimeters (SED);
- ionization chamber dosimeters (ICD).

Passive personal neutron dosimeters — Performance and test requirements

1 Scope

This International Standard provides performance and test requirements for determining the acceptability of personal neutron dosimeters to be used for the measurement of personal dose equivalent, $H_p(10)$, for neutrons ranging in energy from thermal to 20 MeV.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 5-2:2001, *Photography — Density measurements — Part 2: Geometric conditions for transmission density*

ISO 8529-1:2001, *Reference neutron radiations — Part 1: Characteristics and methods of production*

ISO 8529-2:2000, *Reference neutron radiations — Part 2: Calibration fundamentals of radiation protection devices related to the basic quantities characterizing the radiation field*

ISO 8529-3:1998, *Reference neutron radiations — Calibration of area and personal dosimeters and determination of response as a function of energy and angle of incidence*

ISO 12789:2000, *Reference neutron radiations — Characteristics and methods of production of simulated workplace neutron fields*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1 General terms

3.1.1

ageing

changes occurring in an unirradiated dosimeter with time

3.1.2

batch

<detectors or dosimeters> grouping of detectors or dosimeters manufactured according to the same specification and usually at the same time

NOTE 1 Detectors from a batch are intended and supposed to have the same performance characteristics, consistent with the appropriate requirements of this International Standard, and to be used in type or quality test procedures.

NOTE 2 In the case of SSNTD, a batch under this definition can be as small as one single sheet.

3.1.3

detector

device or substance that indicates the presence of a phenomenon without necessarily providing a value of an associated quantity

NOTE In particular, in the presence of radiation, such devices or substances provide by either a direct or indirect means a signal or other indication suitable for measuring one or more quantities.

3.1.4

fading

changes occurring in an irradiated dosimeter with time

3.1.5

personal dosimeter

device comprising one or more detectors positioned in a holder, suitable for the measurement of personal dose equivalent

3.2 Quantities

3.2.1

absorbed dose

D

$d\bar{\varepsilon}$ divided by dm , where $d\bar{\varepsilon}$ is the mean energy imparted by ionizing radiation to matter with mass dm

NOTE 1 The SI unit of absorbed dose is the J/kg. The special name for the unit of absorbed dose is gray (Gy)^[20].

NOTE 2 In quoting values of absorbed dose, it is necessary to specify the material. e.g. soft tissue.

3.2.2

ambient dose equivalent

$H^*(d)$

dose equivalent at a point in a radiation field that would be produced by the corresponding expanded and aligned field in the ICRU sphere at a depth, d , on the radius opposing the direction of the aligned field

NOTE 1 The SI unit of ambient dose equivalent is J/kg. The special name for the unit of ambient dose equivalent is sievert (Sv).

NOTE 2 For strongly penetrating radiation, a depth of 10 mm is currently recommended. The ambient dose equivalent for this depth is then denoted by $H^*(10)$ ^[20].

3.2.3

dose equivalent

H

product of Q and D at a point in soft tissue, where D is the absorbed dose and Q is the quality factor at that point

NOTE The SI unit of dose equivalent is J/kg. The special name for the unit of dose equivalent is sievert (Sv)^[20].

3.2.4

neutron fluence

Φ_n

dN divided by da , where dN is the number of neutrons incident on a sphere of cross-sectional area da

NOTE The SI unit of neutron fluence is m^{-2} , a frequently unit used is cm^{-2} (ISO 8529-1).

3.2.5

personal dose equivalent

$H_p(d)$

dose equivalent in soft tissue, at an appropriate depth, d , below a specified point on the body

NOTE 1 The SI unit of personal dose equivalent is J/kg. The special name for the unit of personal dose equivalent is sievert (Sv)^[20].

NOTE 2 Soft tissue in this context is the ICRU 4-element tissue.

3.2.6

conversion coefficient

$$h_{p\phi}(10;E,\alpha)$$

quotient of the personal dose equivalent, $H_p(10)$, and the neutron fluence, ϕ_n , at a point in the radiation field and used to convert from neutron fluence into the personal dose equivalent at 10 mm depth in the ICRU tissue slab phantom, where E is the energy of the incident neutrons impinging on the phantom at an angle α

NOTE The SI unit of the conversion coefficient is Sv·m². A commonly used unit of the conversion coefficient is pSv·cm².

3.3 Calibration and evaluation

3.3.1

arithmetic mean

$$\bar{x}$$

average of a series of n measurements, x_i , given by the equation

$$\bar{x} = (1/n) \sum_{i=1}^n x_i$$

3.3.2

calibration

set of operations which establish, under a controlled set of standard test conditions, the relationship between the reading given by a dosimeter and the quantity to be measured

3.3.3

calibration factor

$$N$$

quotient of the conventional true value, H_t , divided by the reading, M (see 3.3.16), derived under standard conditions, given by the equation

$$N = \frac{H_t}{M}$$

3.3.4

calibration quantity

physical quantity used to establish the calibration of the dosimeter

NOTE For the purpose of this International Standard, the calibration quantity is the personal dose equivalent at 10 mm depth in the ICRU tissue slab phantom, $H_p(10)$.

3.3.5

coefficient of variation

$$V$$

measure of dispersion for a series of n measurements, x_i , given by the equation

$$V = s/\bar{x}$$

where s is the experimental standard deviation and $\bar{x}$, the arithmetic mean of n measurements.

3.3.6

control specimens

lot of reference detectors or dosimeters of the same type and batch as those used in the test procedures

3.3.7

conventional true value of a quantity

H_t

(dose equivalent) best estimate of a value of a quantity, as determined by a primary or secondary standard or by a reference instrument that has been calibrated against a primary or secondary standard

NOTE A *conventional true value* is regarded as being sufficiently close to the *true value* such that the difference is considered as insignificant for the given purpose.

3.3.8

detection threshold

minimum measured dose equivalent which is significantly higher (at the 95 % confidence level) than the measured dose equivalent of a typical unirradiated dosimeter

3.3.9

experimental standard deviation

s

parameter for a series of n measurements, x_i , characterizing the dispersion and given by the equation

$$s = \sqrt{\frac{1}{n-1} \sum_{i=1}^n (x_i - \bar{x})^2}$$

where $\bar{x}$ is the arithmetic mean of the results of n measurements.

3.3.10

in-field calibration

procedure to calibrate neutron dosimeters in neutron fields representative of a working environment for which the personal dose equivalent rates or neutron spectra and angle distributions have been determined by appropriate methods and hence are sufficiently well known

3.3.11

influence quantity

quantity (parameter) that can have a bearing on the results of a measurement without being the objective of the measurement (ISO 8529-3)

3.3.12

measured dose equivalent

H_M

product of the reading, M , and the calibration factor, N

$$H_M = M \cdot N$$

NOTE In the case of TLAD, elaborate algorithms are generally required (see E.3).

3.3.13

phantom

specified object used to simulate the human body or parts thereof in terms of scattering and absorption properties

NOTE For calibrations, the ISO water slab phantom is employed. It is made with polymethylmethacrylate (PMMA) walls (front wall 2,5 mm thick, other walls 10 mm thick), of outer dimensions 30 cm × 30 cm × 15 cm and filled with water.

3.3.14

quality test

QT

test performed on a number of detectors or dosimeters of a given batch or production designed to assure their quality

NOTE These tests are usually carried out by the user.

3.3.15

read out

process of determining the indication of a detector or dosimeter reader

NOTE For the NTEDs and SSNTs, the reading is usually the number of tracks recorded per unit area; for TLADs, it is the integrated current from the reader; for SEDs, it is the number of bubbles detected or the change in volume; and for ICDs, it is the electric charge collected.

3.3.16

reading

M

quantitative indication of a detector or dosimeter when it is read out, generally corrected for background, ageing, fading and non-linearity of the process or the read out system

3.3.17

reference conditions

set of influence quantities for which the calibration factor is valid without any correction (ISO 8529-3)

3.3.18

response

R

measured dose equivalent, H_M , divided by the conventional true value of the dose equivalent, H_t (see 3.3.7), as given by the following equation

$$R = \frac{H_M}{H_t}$$

NOTE 1 The reading, M , is converted into dose equivalent, H_M , by multiplying M by an appropriate conversion coefficient. In the case of TLAD, elaborate algorithms are generally required (see E.3).

NOTE 2 In this International Standard, the quantity is personal dose equivalent: $R = H_{p,M}(10)/H_{p,t}(10)$.

NOTE 3 In this International Standard for the sake of shortness, $H_M = H$ will be used in Annex C.

NOTE 4 For the specified reference conditions, the response is the reciprocal of the calibration factor.

NOTE 5 In radiation metrology, the term response, abbreviated for this application from “response characteristic”^[15] is defined as the ratio of the reading, M , of the instrument, to the value of the quantity to be measured by the instrument, for a specified type, energy and direction distribution of radiation. It is necessary, in order to avoid confusion, to state the quantity to be measured, e.g. the “fluence response” is the response with respect to the fluence, the “dose equivalent response” is the response with respect to dose equivalent (ISO 8529-3).

3.3.19

standard test conditions

range of values of a set of influence quantities under which a calibration or a determination of response is carried out

3.3.20

type test

TT

performance test of one or more devices that are made to a certain design to show that the design meets given specifications

3.4 Terms with respect to nuclear track emulsion dosimeters (NTED)

3.4.1

film packet

ensemble containing the emulsion wrapped in a light tight envelope that is placed in a sealed pouch, generally filled with dry nitrogen or air, meant to protect the emulsion against fading

3.4.2

latent image

invisible change occurring within the photographic emulsion when it is exposed to actinic radiation, i.e. visible light, ultraviolet or radiation that is directly or indirectly ionizing and that will be converted upon processing into a visible image like a nuclear track

3.4.3

nuclear tracks

tracks created in a nuclear emulsion following the interaction of neutrons with the nuclei in the emulsion, base, wrapping and holder giving rise to protons [by $^{14}\text{N}(\text{n},\text{p})^{14}\text{C}$] or recoil nuclei

3.4.4

nuclear track emulsion

photographic emulsion capable of recording nuclear tracks in a latent form

NOTE These tracks become visible after the chemical development and can be counted under a microscope or any other appropriate scanning device.

3.4.5

nuclear track emulsion dosimeter

passive device consisting of a film packet mounted in a holder (appropriate for the application), intended to be worn on a person's body, detecting neutrons for the purpose of assessing the appropriate personal dose equivalent at or near the position where it is placed

3.4.6

optical density

S

logarithm to the base 10 of the ratio of aperture flux [ISO 5-2] to the flux transmitted by the sample under the same beam geometric conditions

3.4.7

scanning

process of evaluating a nuclear track emulsion by counting the visible tracks under a microscope either by an operator or by an automatic scanning device

3.4.8

stability of latent image

degree to which a nuclear track emulsion is capable of producing a developed image of a nuclear interaction irrespective of the time elapsed between the formation of the latent image and the development of the emulsion and irrespective of the ambient conditions that have prevailed during this time (temperature or humidity)

3.4.9

track density

number of tracks scanned per unit area

3.5 Terms with respect to solid state nuclear track dosimeters (SSNTD)

3.5.1

chemical etch bath

configuration containing the etching solution for etching detectors under a defined temperature

3.5.2

converter material

material in which neutrons undergo nuclear reactions and produce charged particles that can be detected by SSNT

NOTE Examples are hydrogen compounds for fast neutrons, and ^{10}B , ^6Li or nitrogen-containing compounds for thermal and epithermal neutrons.

3.5.3

etch chamber

device containing the detectors in a geometry allowing their proper etching under defined temperature conditions and an applied voltage (if applicable)

3.5.4

etching

process to develop the radiation induced tracks in detectors to make them countable

NOTE Etching could be chemical by putting the detectors into a proper chemical solution under defined temperature conditions. When an alternating high voltage is applied across the detector, then etching is called electro-chemical.

3.5.5

solid state nuclear track dosimeter

passive device consisting of one or more track etch detectors mounted in a holder (appropriate for the application), intended to be worn on a person's body to detect neutrons for the purpose of assessing the appropriate personal dose equivalent at or near the position where it is placed

3.5.6

track etch detector

material, usually plastic in nature, carefully manufactured under controlled conditions for the purpose of radiation measurements

3.5.7

track etch reader

device used to establish the number of tracks per unit area

3.6 Terms with respect to thermoluminescence albedo dosimeters (TLAD)

3.6.1

albedo

fraction of incident radiation scattered from a surface

NOTE Neutrons scattered back from a body following interactions within the body are called albedo neutrons.

3.6.2

annealing

controlled thermal treatment of a TL detector or dosimeter during or after readout

3.6.3

apparent photon dose equivalent

H_a

measured dose equivalent of each detector evaluated as if it had been irradiated by reference photon radiation

3.6.4

preparation

normal treatment of annealing, cleaning, etc., of detectors or dosimeters that are intended for routine use

3.6.5

thermoluminescence

TL

emission of light exhibited by certain substances or materials when the substance is heated following its exposure to ionizing radiation or U.V.

NOTE Strictly, the property should be referred to as radiothermoluminescence, but the abbreviated form is usually adequate.

3.6.6

thermoluminescence albedo neutron dosimeter

albedo dosimeter

passive device, consisting of two or more TL detectors, mounted in a holder (appropriate for the application), intended to detect incident and albedo neutrons when worn on a person's body for the purpose of assessing the appropriate personal dose equivalent at or near the position where it is placed

3.6.7

thermoluminescent detector

specified quantity of TL material, or such material incorporated with other non-luminescent material into a matrix, being defined by mass, shape or size or the mass of material incorporated in the matrix

3.6.8

thermoluminescence dosimeter reader

TL reader

instrument used to measure the light emitted from the detectors of thermoluminescence dosimeters, consisting essentially of a heating device, a light measuring device and the associated electronics

3.6.9

zero point

reading of unirradiated TLDs expressed in apparent photon dose equivalent

3.7 Terms with respect to superheated emulsion dosimeters (SED)

3.7.1

activation

procedure to render superheated emulsions ready for use

NOTE 1 Bubble detectors (BD) are typically stored and kept inactive under pressure, which is applied keeping their screw caps on. They are activated by removing the screw cap.

NOTE 2 Superheated drop detectors (SDD) are typically stored and kept inactive at reduced temperature. They are activated by allowing them to reach ambient temperature.

3.7.2

bubble reading

reading of a dosimeter based on superheated emulsions

NOTE In bubble detectors (BD), the number of visible bubbles, that comprise the measurement, is obtained by optical counting or by use of an automatic electro-optical instrument. In the superheated drop detectors, the volume of evolved gas is measured with the help of a calibrated scale.

3.7.3

depletion

decrease in the number of superheated drops due to their transformation into bubbles after being struck by neutrons

NOTE In the bubble detectors (BD), the initial sensitivity is restored by repressurization and no depletion correction is necessary. In the superheated drop detectors, the bubbles separate permanently from the detector and a depletion correction is necessary with prolonged use.

3.7.4

resetting

clearing of bubbles from the visco-elastic matrix

NOTE In the bubble detectors (BD), this is done by the application of external pressure using a piston assembly. In the superheated drop detectors, resetting is unnecessary because the bubbles leave the medium by rising to the top of the dosimeter due to their buoyancy.

3.7.5

superheated drops

drops of superheated liquid, generally of 5 µm to 100 µm diameter

NOTE These drops dispersed throughout a visco-elastic matrix represent the radiation-sensitive sites of the detector.

3.7.6

superheated emulsion

SE

superheated drops dispersed within a clear visco-elastic matrix

NOTE Superheated emulsions are typically contained in a device consisting of a transparent, cylindrical holder. Visible bubbles are formed throughout the device when irradiated with neutrons. The number of bubbles provides a measure of the neutron dose.

3.7.7

superheated liquid

fluid above its normal boiling point that is still in the liquid phase

3.7.8

visco-elastic matrix

(gel) aqueous or polymer gel in which immiscible superheated drops are uniformly emulsified

NOTE In bubble detectors, the visco-elastic matrix typically keeps bubbles immobilized at the location of their formation. In superheated drop detectors (SDD), the visco-elastic matrix typically allows bubbles to rise out of the detector by buoyancy.

3.8 Terms with respect to ion chamber dosimeters with direct ion storage (ICD)

3.8.1

apparent photon dose equivalent

H_a

measured dose equivalent of each detector evaluated as if it had been irradiated by reference photon radiation

3.8.2

direct ion storage

DIS

permanent storage of an electric charge in a MOSFET with an open floating gate connected to an ion chamber

3.8.3

ion chamber detector

air volume surrounded by conductive wall material with applied electric field to collect ions produced by ionizing radiation

NOTE The neutron sensitivity of an ion chamber strongly depends on the wall material. Pairs of ion chambers with different neutron sensitivities are used to differentiate between neutron and photon radiation.

4 Units

This International Standard uses SI units. However the following units of practical importance for time are used where necessary: days (d) and hours (h).

5 General test conditions

5.1 Test conditions

All tests shall be performed under standard test conditions (see Annex A), except where otherwise stated. The actual conditions should be indicated in the test report. They should not undergo large or rapid changes during a series of measurements.

Irradiation of dosimeters shall be performed on a phantom in terms of personal dose equivalent for the energy and angle dependence of response tests. For the other tests, dosimeters may be irradiated free-in-air and reference quantity could be neutron fluence.

5.2 Reference radiation

The reference radiation fields defined in ISO 8529-1 shall be used. In addition, simulated workplace neutron fields may be used (ISO 12789). For many of the tests, it is sufficient to use only one neutron source (e.g. ^{241}Am -Be or ^{252}Cf neutron sources). Information on irradiation conditions is found in Annex D.

NOTE For TLADs, any neutron source with adequate low-energy neutron intensity, the more moderated the better, can be used for the tests.

5.3 Test requirements

All detectors and dosimeters shall be subjected to the preparation, handling and read-out recommended by the manufacturer. Packaging, handling and processing of the detectors and dosimeters govern their performance that is generally assessed under laboratory conditions. As these conditions can never adequately simulate the conditions actually experienced in practical personal dosimetry, caution is therefore necessary when applying the results of these performance tests in real situations.

6 Performance requirements

The following general requirements apply to the tests for all detectors and dosimeters.

- a) The tests are performed on a specified number of dosimeters randomly selected from one batch.
- b) Type tests are made to assess the basic characteristics of the detectors or dosimeters and are often assured by recognized national laboratories, while quality tests are usually done by dosimetry services that intend to use dosimeters for routine neutron personal dosimetry.
- c) The performance requirements are listed in Tables 6.1 to 6.5 as type tests (TT) or quality tests (QT).
- d) The number of dosimeters used in each test shall be such to demonstrate that the requirements are met at the 95 % confidence level.
- e) The reader used in the evaluation of the detectors and dosimeters shall be subjected to routine tests to establish consistency of operation.

Information on the characteristics of detectors and dosimeters is given in Annex E.

Table 1 — Performance requirements for nuclear track emulsion dosimeters

No.	Performance characteristics	Performance requirements	Test
1	Batch homogeneity	The coefficient of variation of the reading for n dosimeters shall not exceed 30 % for a given personal dose equivalent greater than 1 mSv.	QT C.1.1
2	Reproducibility	Not applicable	—
3	Linearity	The response to neutrons shall not vary by more than 20 % over the personal dose equivalent range from 0,5 mSv to 20 mSv.	TT C.1.2
4	Detection threshold	The detection threshold shall not exceed a personal dose equivalent of 0,2 mSv.	QT C.1.3
5	Fading (stability of the latent image)	The fading of tracks under normal test conditions shall not be greater than 10 %, 15 %, 20 % and 30 % for periods of 3 days, 10 days, 30 days and 90 days, respectively, when comparing the test samples with control specimens.	QT C.1.4
6	Ageing	Not applicable	—
7	Residual signal	Not applicable	—
8	Self-irradiation	Not applicable	—
9	Exposure to radiation other than neutrons	The arithmetic mean of the response to neutrons shall differ by less than 10 % for a personal dose equivalent between 1 mSv and 2 mSv and an additional exposure to 3 mSv of photon radiation from a ^{137}Cs source compared with dosimeters not subjected to photon radiation.	TT C.1.5
10	Stability under various climatic conditions	The fading of tracks at a temperature of 40 °C and a humidity of 100 % shall not be greater than 15 %, 30 % and 50 % for periods of 3 days, 10 days and 30 days when comparing the test samples with control specimens.	TT C.1.6
11	Effect of light exposure (opacity to light)	No streaks shall be observed on the processed test film when exposed to 1 000 W/m ² for one hour using a xenon lamp equivalent to bright sunlight (295 nm to 769 nm). The mean optical density over the useful area of the film shall not differ by more than three experimental standard deviations in comparison with the mean optical density of the control specimen.	TT C.1.7
12	Physical damage	Not applicable	—
13	Energy dependence of response	The response at normal incidence in the stated energy range for the dosimetry system shall not vary by more than ± 50 % for a personal dose equivalent of a least 1 mSv.	TT C.1.8
14	Angle dependence of response	The arithmetic mean of the response of a dosimeter at angles of incidence of 0°, 15°, 45° and 60° from normal shall not differ by more than 30 % from the corresponding response at normal incidence.	TT C.1.9

Table 2 — Performance requirements for solid state nuclear track dosimeters

No.	Performance characteristics	Performance requirements	Test
1	Batch homogeneity	The coefficient of variation of the reading for n dosimeters shall not exceed 15 % for any given personal dose equivalent greater than 1 mSv.	QT C.2.1
2	Reproducibility	Not applicable	—
3	Linearity	The response to neutrons shall not vary by more than 10 % over the personal dose equivalent range from 1 mSv to 10 mSv.	TT C.2.2
4	Detection threshold	The detection threshold shall not exceed a personal dose equivalent of 0,3 mSv.	QT C.2.3
5	Fading	The response of dosimeters irradiated at the beginning of a storage period shall not vary by more than 30 % for a 90-day storage period under standard test conditions. (The period from the manufacture of the detectors to the beginning of the test shall not exceed 90 days.)	TT C.2.4
6	Ageing	The response of dosimeters irradiated at the end of a storage period shall not vary by more than 30 % for a 90-day storage period under standard test conditions. (The period from the manufacture of the detectors to the beginning of the test shall not exceed 90 days.)	TT C.2.5
7	Residual signal	Not applicable	—
8	Self-irradiation	Not applicable	—
9	Exposure to radiation other than neutrons	The response to radiation other than neutrons, in particular ^{137}Cs photons, shall be less than 10 % in terms of $H_p(10)$. The zero point shall not change by more than 0,5 mSv for dosimeters exposed to 3 MBq · h/m ³ of radon at 50 % equilibrium with daughters ($F = 0,5$).	TT C.2.6
10	Stability under various climatic conditions	The measured dose equivalent shall not differ from the conventional true value by more than 20 % for 48-h storage at 40 °C and 90 % relative humidity when the dosimeters are irradiated either at the beginning or at the end of the storage period.	TT C.2.7
11	Effect of light exposure	The zero point shall not change by more than 1 mSv when exposed to a xenon lamp equivalent to bright sunlight (295 nm to 769 nm) to 1 000 W/m ² for one day. For exposure of one week, the reading shall not differ from the reading of a dosimeter kept in the dark by more than 10 %.	TT C.2.8
12	Physical damage	Not applicable	—
13	Energy dependence of response	The response at normal incidence in the stated energy range for the dosimetry system shall not vary by more than ± 50 % for a personal dose equivalent of at least 1 mSv.	TT C.2.9
14	Angle dependence of response	The arithmetic mean of the response of a dosimeter at angles of incidence of 0°, 15°, 45° and 60° from normal shall not differ by more than 40 % from the corresponding response at normal incidence. The personal dose equivalent should be at least 1 mSv.	TT C.2.10

Table 3 — Performance requirements for thermoluminescence albedo dosimeters

No.	Performance characteristics	Performance requirements	Test
1	Batch homogeneity	The coefficient of variation of the apparent photon dose equivalent for n dosimeters shall not exceed 20 %.	QT C.3.1
2	Reproducibility	The coefficient of variation of the apparent photon dose equivalent for n dosimeters shall not exceed 20 % for each dosimeter separately when irradiated ten times.	QT C.3.2
3	Linearity	The response shall not vary by more than 10 % when dosimeters are irradiated with neutrons and photons producing readings of apparent photon dose equivalents between 1 mSv and 100 mSv.	TT C.3.3
4	Detection threshold	The detection threshold shall not exceed an apparent photon dose equivalent of 0,3 mSv.	QT C.3.4
5	Fading	The response of dosimeters irradiated at the beginning of a storage period shall not change by more than 10 % for a 90-day storage period under standard test conditions.	TT C.3.5
6	Ageing	Not applicable	—
7	Residual signal	The residual signal shall not exceed an apparent photon dose equivalent to 0,3 mSv following an irradiation corresponding to an apparent photon dose equivalent to 10 mSv.	TT C.3.6
8	Self-irradiation	After a storage period of 60 days, the zero point shall not change by an apparent photon dose equivalent of more than 0,3 mSv.	TT C.3.7
9	Exposure to radiation other than neutrons	Not applicable (see E.3)	—
10	Stability under various climatic conditions	The apparent photon dose equivalent shall not differ from the conventional true value by more than 5 % after 30 days storage under standard test conditions or 10 % for 48-h storage at 40 °C and 90 % relative humidity, when the dosimeters are irradiated either at the beginning or at the end of the storage period.	TT C.3.8
11	Effect of light exposure	The zero point shall not change by more than an apparent photon dose equivalent of 0,3 mSv when exposed with a xenon lamp equivalent to bright sunlight (295 nm to 769 nm) to 1 000 W/m ² for one day. For exposure of one week, the apparent photon dose equivalent shall not differ from the apparent photon dose equivalent of a dosimeter kept in the dark by more than 10 %.	TT C.3.9
12	Physical damage	Not applicable	—
13	Energy dependence of response	Not applicable (see E.3)	—
14	Angle dependence of response	The arithmetic mean of the response of a dosimeter at angles of incidence of 0°, 15°, 45° and 60° from normal shall not differ by more than 30 % from the corresponding response at normal incidence. The personal dose equivalent should be about 1 mSv.	TT C.3.10

Table 4 — Performance requirements for superheated emulsions

No.	Performance characteristics	Performance requirements	Test
1	Batch homogeneity	The coefficient of variation of the reading for n dosimeters shall not exceed 25 % for a personal dose equivalent greater than 0,1 mSv.	QT C.4.1
2	Reproducibility	The coefficient of variation of the reading for n irradiations shall not exceed 25 % for any given personal dose equivalent less than 1 mSv.	QT C.4.2
3	Linearity	The response to neutrons shall not vary by more than 25 % over the personal dose equivalent range from 0,1 mSv to 1 mSv.	TT C.4.3
4	Detection threshold	The detection threshold shall not exceed a personal dose equivalent of 0,1 mSv.	QT C.4.4
5	Fading	Not applicable	—
6	Ageing	Bubble detectors shall not change in response by more than 25 % over up to 3 months, if used and reset and stored in an inactivated state following daily use.	TT C.4.5
7	Residual signal	Following an irradiation of up to 1 mSv, no bubble should be present in the detector after the resetting procedure.	QT C.4.6
8	Self-irradiation	Not applicable	—
9	Exposure to radiation other than neutrons	Not applicable	—
10	Stability under various climatic conditions	The response shall not vary by more than 25 % over the temperature range from 20 °C to 40 °C.	TT C.4.7
11	Effect of light exposure	Not applicable	—
12	Physical damage	A drop from a height of 1 m should not produce more than four bubbles.	QT C.4.8
13	Energy dependence of response	The response shall not vary by more than 25 % upon irradiation with mono-energetic neutrons (ISO 8529-1) over the energy range specified for the detector.	TT C.4.9
14	Angle dependence of response	The arithmetic mean of the response of a dosimeter at angles of incidence of 0°, 15°, 45° and 60° from normal shall not differ by more than 30 % from the corresponding response at normal incidence. The personal dose equivalent should be such that at least 100 bubbles are formed in a detector.	QT C.4.10

Table 5 — Performance requirements for ion chamber detectors with direct ion storage

No.	Performance characteristics	Performance requirements	Test
1	Batch homogeneity	The coefficient of variation of the apparent photon dose equivalent for n dosimeters shall not exceed 20 %.	QT C.5.1
2	Reproducibility	The coefficient of variation of the apparent photon dose equivalent for n dosimeters shall not exceed 20 % for each dosimeter separately when irradiated ten times.	QT C.5.2
3	Linearity	The response shall not vary by more than 10 % when dosimeters are irradiated with neutrons and photons producing readings corresponding to apparent photon dose equivalents between 1 mSv and 100 mSv.	TT C.5.3
4	Detection threshold	The detection threshold shall not exceed an apparent photon dose equivalent of 0,3 mSv.	QT C.5.4
5	Fading	The response of dosimeters irradiated at the beginning of a storage period shall not vary by more than 10 % for a 90-day storage period under standard test conditions.	TT C.5.5
6	Ageing	The response of dosimeters irradiated at the end of a storage period shall not vary by more than 10 % for a 90-day storage period under standard test conditions.	TT C.5.6
7	Residual signal	Not applicable	—
8	Self-irradiation	Not applicable	—
9	Exposure to radiation other than neutrons	The arithmetic mean of the response to neutrons shall differ by less than 20 % for a personal dose equivalent between 1 mSv and 2 mSv and an additional exposure to 3 mSv of photon radiation from a ^{137}Cs source compared with dosimeters not subjected to photon radiation.	TT C.5.7
10	Stability under various climatic conditions	The apparent photon dose equivalent shall not differ from the conventional true value by more than 10 % for 48-h storage at 40 °C and 90 % relative humidity when the dosimeters are irradiated either at the beginning or at the end of the storage period.	TT C.5.8
11	Effect of light exposure	Not applicable	—
12	Physical damage	Not applicable	—
13	Energy dependence of response	The response at normal incidence in the stated energy range for the dosimetry system shall not vary by more than ± 50 % for a personal dose equivalent of at least 1 mSv.	TT C.5.9
14	Angle dependence of response	The arithmetic mean of the response of a dosimeter at angles of incidence of 0°, 15°, 45° and 60° from normal shall not differ by more than 30 % from the corresponding response at normal incidence. The personal dose equivalent should be at least 1 mSv.	TT C.5.10

7 Test methods

Tests to demonstrate compliance with the specified performance requirements are detailed in Annex C. These provide guidance; specific details pertaining to similar tests by various national regulatory agencies shall take precedence over those described here.

8 Identification and accompanying documentation

8.1 Individual marking

Dosemeters and detectors shall have simple, unique and secure means of identification. The marking shall not impair the useful portion of the detector either directly or indirectly, nor shall it change its behaviour in any significant manner. Dosemeters shall carry any necessary markings for determining their origin, expiry date, if relevant, and that they are intended for neutron dosimetry.

8.2 Collective marking

The following information shall be indicated on each box (or other collective packing) of detectors or doseimeters or, failing this, on an accompanying note:

- name or trademark of the manufacturer;
- complete designation;
- series number or manufacturer's batch number;
- expiry date, if relevant.

8.3 Accompanying documentation

The note attached to each box or other packing container shall carry at least the following information, if it is not provided on the container:

- complete designation;
- name and trademark of manufacturer;
- neutron energy range for which the detectors or doseimeters are designed;
- method of processing.

9 Certification

An acknowledged testing laboratory will make the certification. Some basic requirements are suggested for manufacturers claiming compliance (or partial compliance) with the performance requirements of this International Standard:

- a) name and address of the manufacturer;
- b) name and address of the testing laboratory or laboratories, when more than one is involved;
- c) description/specification of the doseimeters under consideration, including how the doseimeter is assembled; drawings or photos are encouraged;
- d) description of reader, procedures and ancillary equipment used (model No., etc.);
- e) statements giving the dates and the outcome of each test (test reports shall be available for inspection by the purchaser);
- f) details of conversion coefficients and phantoms agreed between the manufacturer and the purchaser, when those proposed in this International Standard are not used;
- g) any deviations from the recommended testing procedures should be justified.

Annex A (normative)

Reference and standard test conditions

Table A.1

Influence	Reference conditions (unless otherwise stated by the manufacturer)	Standard test conditions (unless otherwise stated by the manufacturer)
Reference neutron radiation	Conforming to ISO 8529-3	Conforming to ISO 8529-3 and ISO 12789
Ambient temperature	$T = 20\text{ °C}$	$T = 18\text{ °C}$ to 22 °C
Relative humidity	65 %	50 % to 75 %
Atmospheric pressure	$p = 101,3\text{ kPa}$	$p = 86$ to 106 kPa
Radiation background	$H^*(10) \leq 0,1\text{ }\mu\text{Sv/h}$	$H^*(10) \leq 0,25\text{ }\mu\text{Sv/h}$
Contamination by radioactive elements	Negligible	Negligible
Light intensity	50 W/m^2	$< 100\text{ W/m}^2$


Annex B (normative)

Confidence limits

B.1 General

The magnitude of the uncertainty of a measured value due to random effects is a significant fraction of the allowed tolerances of this measured value. This uncertainty is considered by taking more than one measurement.

In fact, the number of measurements should be large enough to ensure that the arithmetic mean, $\bar{x}$, of a series of measurements provides a reliable estimate of the “true” mean, μ , and that in addition the experimental standard deviation, s , is a reliable estimate of the standard deviation, σ ^[10]. In this International Standard, the number of measurements or the sample size shall be chosen in such a way that, for a confidence level of 95 %, the confidence interval obtained for each arithmetic mean, $\bar{x}$, lies either within the limits of variation of the measured value allowed in the test (test passed, point Δ in Figure B.1) or outside of these limits (test failed, point $\circ$ in Figure B.1). If one of the allowed limits of variation, x_u or x_l , lies within the confidence interval of the arithmetic mean (points $\square$ in Figure B.1), the number of measurements or the sample size can be increased to reduce the width $2l$ of the confidence interval of the arithmetic mean, $\bar{x}$, in order to reach one of the two cases mentioned above that are necessary for an unequivocal decision of passing the test or not.



Key

- 1 confidence interval of the arithmetic mean $\bar{x}$ of width $2l$
- 2 allowed upper limit of variation, x_u
- 3 allowed lower limit of variation, x_l

Figure B.1 — Test for confidence interval

The test is passed if the confidence interval of width $2l$ around $\bar{x}$ lies between the allowed upper and lower limits of variation x_u and x_l as given by Equation (B.1):

$$x_l + l < \bar{x} < x_u - l \quad (\text{B.1})$$

In each test, it is recommended to start with 10 measurements and to increase the number of measurements or doseimeters if it turns out to be necessary to reduce the width $2l$ of the confidence interval.

B.2 Confidence interval for the arithmetic mean, $\bar{x}$

The confidence interval for the arithmetic mean, $\bar{x}$, is $(\bar{x} - l_i, \bar{x} + l_i)$, where l_i is the half width of the confidence interval of $\bar{x}$ relative to the i th series of measurements. When calculating $\bar{x}$ from n_i measurements, the half width of the confidence interval is given by Equation (B.2):

$$l_i = t_n \cdot s_i / \sqrt{n_i} \quad (\text{B.2})$$

where

s_i is the experimental standard deviation for the i th series of measurements;

t_n is taken from Table B.1 for n_i measurements.

EXAMPLE For $n_i = 10$; $l_i = 2,26 \cdot s_i / \sqrt{10} = 0,71 \cdot s_i$.

Table B.1 — Two-sided Student's t values for 95 % confidence interval for sample size n_i

n_i	t_n	n_i	t_n
2	12,71	15	2,15
3	4,30	20	2,09
4	3,18	25	2,06
5	2,78	30	2,05
6	2,57	40	2,02
7	2,45	50	2,00
8	2,37	60	1,98
9	2,31	120	1,96
10	2,26	∞	1,96

B.3 Confidence interval for the experimental standard deviation, s

The confidence interval for the experimental standard deviation, s , of the arithmetic mean, $\bar{x}$, is given by the expressions

$$(s - l_s, s + l_s)$$

where l_s is the half width of the confidence interval of s . If s is calculated from n measurements for which the probability distribution is close to normal, the upper limit of l_s at a confidence level of 95 % is approximated by Equation (B.3):

$$l_s(n) = t_n \cdot s \sqrt{0,5/(n-1)} \quad (\text{B.3})$$

EXAMPLE $l_s(10) = 0,53 \cdot s$ is obtained for $n = 10$ measurements and t_n is taken from Table B.1.

B.4 Confidence interval for a combined quantity

If the limits of variation are stated for a quantity, the arithmetic mean, $\bar{x}$, calculated from k means $\bar{x}_1, \bar{x}_2, \bar{x}_3, \dots, \bar{x}_k$ is given by Equation (B.4):

$$\bar{x} = f(\bar{x}_1, \bar{x}_2, \bar{x}_3, \dots, \bar{x}_k) \quad (\text{B.4})$$

and the half width of the confidence interval of the i th arithmetic mean is l_i . The half width, l , of the confidence interval for $\bar{x}$ is given by Equation (B.5):

$$l = \sqrt{\sum_{i=1}^k \left[\delta f(\bar{x}_1, \bar{x}_2, \bar{x}_3, \dots, \bar{x}_k) \delta x_i \right]^2 \cdot l_i^2} \quad (\text{B.5})$$

EXAMPLES

$$\text{a) } \bar{x} = \bar{x}_1 \pm \bar{x}_2 \quad \text{hence} \quad l = \sqrt{l_1^2 + l_2^2}$$

$$\text{In general, } \bar{x} = \sum_{i=1}^n \bar{x}_i \quad \text{hence} \quad l = \sqrt{\sum_{i=1}^n l_i^2}$$

$$\text{b) } \bar{x} = \bar{x}_1 / \bar{x}_2 \quad \text{hence} \quad l = (\bar{x}_1 / \bar{x}_2) \sqrt{(l_1 / \bar{x}_1)^2 + (l_2 / \bar{x}_2)^2}$$

$$\text{c) } \bar{x} = (\bar{x}_1 - \bar{x}_2) / (\bar{x}_1 + \bar{x}_2) \quad \text{hence} \quad l = \frac{2}{(\bar{x}_1 + \bar{x}_2)^2} \sqrt{(\bar{x}_2 \cdot l_1)^2 + (\bar{x}_1 \cdot l_2)^2}$$

B.5 Confidence interval for detection threshold

In the case for the calculation of the detection threshold according to the following formula (used elsewhere to define the critical level of the decision limit^[27]):

$$t_n \cdot s \quad (\text{B.6})$$

The Student's t values to be used are the ones reported in the following table belonging to the one-sided Student's t distribution.

Table B.2 — One-sided Student's t values for 95 % confidence interval for sample size n_i

n_i	t_n	n_i	t_n
2	6,32	15	1,76
3	2,92	20	1,73
4	2,35	25	1,71
5	2,13	30	1,70
6	2,01	40	1,68
7	1,94	50	1,68
8	1,89	60	1,67
9	1,86	120	1,66
10	1,83	∞	1,64

Annex C (normative)

Performance tests

C.1 Nuclear track emulsions (NTE)

C.1.1 Batch homogeneity

Prepare and irradiate a lot of $i = 1$ to n dosimeters of the same emulsion with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to the same conventional true value of the personal dose equivalent, H_t , of at least 1 mSv.

Develop, read out (scan at least 100 tracks), determine the track density for each dosimeter, M_i , and show that the coefficient of variation, V , for the reading, M_i , of the n dosimeters does not exceed 30 %.

C.1.2 Linearity

Prepare six lots numbered $j = 1$ to 6 of $i = 1$ to n dosimeters and irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to a conventional true value of personal dose equivalents, $H_{t,j}$, of 0,5 mSv, 1 mSv, 2 mSv, 5 mSv, 10 mSv and 20 mSv, respectively.

Develop, read out (scan at least 100 tracks) each dosimeter, determine the measured dose equivalent, $H_{i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each lot j and the respective experimental standard deviations, s_j . Show that for each personal dose equivalent, H_t meets the criterion given in Equation (C.1):

$$0,8 \leq (\bar{H}_j \pm l_{i,j}) / H_{t,j} \leq 1,2 \quad (\text{C.1})$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.1.3 Detection threshold

Develop and read out (scan) n unirradiated dosimeters. Use the conversion coefficient for a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to transform the number of tracks into measured dose equivalent, H_t , and calculate the experimental standard deviation, s , for the n dosimeters. Show that the criterion in Equation (C.2) is met:

$$t_n \cdot s \leq 0,2 \text{ mSv}^{[27]} \quad (\text{C.2})$$

where the values for t_n are listed in Table B.2.

C.1.4 Fading (stability of the latent image)

Prepare five lots numbered $j = 1$ to 5 of $n + 2$ dosimeters. Irradiate $i = 1$ to n dosimeters of each lot with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to the same value of the conventional true value of the personal dose equivalent, H_t , between 1 mSv and 2 mSv at days D_0 (lot 1), $D_0 + 60$ (lot 2), $D_0 + 80$ (lot 3), $D_0 + 87$ (lot 4) and $D_0 + 90$ (lot 5) and keep two dosimeters of each lot as background samples. Following irradiation, store and keep all dosimeters under normal test conditions. Develop all dosimeters together on days $D_0 + 91$ or $D_0 + 92$ and consider the dosimeters of lot five as the control specimens.

Read out (scan at least 100 tracks) each dosimeter, determine the track density, M_{ij} , and check the background dosimeters for an undue increase of tracks. Determine the arithmetic mean of the reading, $\bar{M}_j$, and the experimental standard deviation, s_j , for each of the five lots. Show that the criterion of Equation (C.3) is met:

$$\left(\bar{M}_j / \bar{M}_5\right) \pm l \geq Q_{R,j} \quad (\text{C.3})$$

where $j = 1$ to 4, the Q_{R1} to Q_{R4} are 0,7, 0,8, 0,85 and 0,9, respectively, and l for the quotient of the arithmetic means is calculated in accordance with Equation (B.5).

C.1.5 Exposure to radiation other than neutrons

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters and irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to a conventional true value of the personal dose equivalent, H_t , between 1 mSv and 2 mSv. In addition irradiate the second lot with 3 mSv of photon radiation from a ^{137}Cs source.

Develop, read out (scan at least 100 tracks) and determine the track density for each dosimeter, M_{ij} . Determine the arithmetic mean of the reading, $\bar{M}_j$, and the experimental standard deviation, s_j , for the two lots. Show that the criterion in Equation (C.4) is met:

$$0,9 \leq \left(\bar{M}_1 / \bar{M}_2\right) \pm l \leq 1,1 \quad (\text{C.4})$$

where l for the quotient of the arithmetic means is calculated in accordance with Equation (B.5).

C.1.6 Stability under various climatic conditions

Prepare four lots of $n + 2$ dosimeters each and identify the lots by numbers $j = 1$ to 4. Irradiate $i = 1$ to n dosimeters of each lot with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to the same conventional true value of the personal dose equivalent, H_t , between 1 mSv and 2 mSv at days D_0 (lot 1), $D_0 + 20$ (lot 2), $D_0 + 27$ (lot 3) and $D_0 + 30$ (lot 4) and keep two dosimeters of each lot as background samples. Following irradiation, store and keep all dosimeters at 40 °C and a relative humidity of 100 %.

Develop all dosimeters together on days $D_0 + 31$ or $D_0 + 32$ and consider the dosimeters of lot 4 as the control specimens. Read out (scan at least 100 tracks) and determine the track density for each dosimeter, M_{ij} , and check the background dosimeters for an undue increase of tracks in the samples. Determine the arithmetic mean of the reading, $\bar{M}_j$, and the experimental standard deviation, s_j , for each of the four lots. Show that the criterion in Equation (C.5) is met:

$$\left(\bar{M}_j / \bar{M}_4\right) \pm l \geq Q_{R,j} \quad (\text{C.5})$$

where the Q_{R1} to Q_{R3} are 0,5, 0,7 and 0,85, respectively, and l for the quotient of the arithmetic means is calculated in accordance with Equation (B.5).

C.1.7 Effect of light exposure (opacity to light)

Expose film packets randomly selected from a batch to an illumination from a xenon lamp of 1 000 W/m² for 1 h at a sufficiently large distance from the light source to avoid the film packets exceeding a temperature of 30 °C. Set the face of the film packets perpendicular to the luminous radiation. Then expose the other face of the film packets under the same conditions for 1 h. Then expose each edge of the film packets under the same conditions for 1 h.

Process films simultaneously with an unexposed control specimen and compare their optical densities with respect to streaks or lack of homogeneity of the optical density. The optical density, $S_{k,j}$, is measured at ten different points on each film j . Calculate the mean optical density, $\bar{S}_j$, and the experimental standard deviation, s_j , for each film and compare the values with those of the control specimen, S_c . Show that the criterion in Equation (C.6) is met:

$$\left| (\bar{S}_j - \bar{S}_c) \right| + l \leq 3 \cdot s_j \quad (\text{C.6})$$

where l for the difference of the arithmetic means is calculated in accordance with Equation (B.5).

C.1.8 Energy dependence of response

Prepare m lots numbered $j = 1$ to m of $i = 1$ to n dosimeters and irradiate them with m neutron energies in the applicable energy range for which compliance with this standard is sought. These neutron energies should be taken from ISO 8529-1:2001, Table 2. The dosimeters shall be irradiated on the ISO slab water phantom to a conventional true value of the personal dose equivalent, H_t , between 1 mSv and 2 mSv.

Develop, read out (scan at least 100 tracks), determine the measured dose equivalent for each dosimeter, $H_{i,j}$, and calculate for each lot j (energy), the arithmetic mean, $\bar{H}_j$, and the respective experimental standard deviation, s_j . Show that the criterion in Equation (C.7) is met:

$$0,5 \leq (\bar{H}_j \pm l_{i,j}) / H_{t,j} \leq 1,5 \quad (\text{C.7})$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.1.9 Angle dependence of response

Prepare four lots numbered $j = 1$ to 4 of four dosimeters each numbered $i = 1$ to 4. Irradiate all dosimeters with neutrons to a value of the conventional true value of the personal dose equivalent, H_t , between 1 and 2 mSv on the ISO slab water phantom under the following conditions (ISO 8529-3):

- lot 1: normal incidence;
- lot 2: 15° off normal;
- lot 3: 45° off normal;
- lot 4: 60° off normal.

For lots 2 to 4, the angle of incidence for the dosimeters $i = 1$ to 4 irradiated is varied positive and negative in two planes perpendicular to each other and to the plane of the dosimeter. Determine the conventional true value of the personal dose equivalent employing the appropriate conversion coefficient for the neutrons used in the test at the respective angles j , H_t .

Develop, read out (scan at least 100 tracks), determine the measured dose equivalent for each dosimeter, $H_{i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (angles), the respective experimental standard deviation, s_j , and the response, R_j . Show that the criterion in Equation (C.8) is met:

$$0,7 \leq \left[0,25 \sum_{j=1}^4 (R_j / R_1) \right] \pm l \leq 1,3 \quad (\text{C.8})$$

where l for the sum of the ratios is calculated in accordance with Equation (B.5).

C.2 Solid state nuclear track detectors (SSNT)

C.2.1 Batch homogeneity

Prepare and irradiate a lot of $i = 1$ to n dosimeters belonging to the same batch (sheet or material) with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to the same conventional true value of the personal dose equivalent, H_t , of at least 1 mSv.

Etch, read out each dosimeter and show that the coefficient of variation, V , for the reading, M_i , of n dosimeters does not exceed 15 %.

C.2.2 Linearity

Prepare five lots numbered $j = 1$ to 5 of $i = 1$ to n dosimeters each and irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to a conventional true value of personal dose equivalents, $H_{t,j}$, of 1 mSv, 2 mSv, 5 mSv, 10 mSv and 20 mSv, respectively.

Etch, read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each lot j and the respective experimental standard deviations, s_j . Show that for each personal dose equivalent, $H_{t,j}$, the criterion in Equation (C.9) is met:

$$0,85 \leq (\bar{H}_j \pm l_{i,j}) / H_{t,j} \leq 1,15 \quad (\text{C.9})$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.2.3 Detection threshold

Etch and read out $i = 1$ to n unirradiated dosimeters. Use the conversion coefficient for a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to transform the number of tracks into measured dose equivalent, H_i , and calculate the experimental standard deviation, s , for the n dosimeters. Show that the criterion, expressed in millisieverts, in Equation (C.10) is met:

$$t_n \cdot s \leq 0,3^{[27]} \quad (\text{C.10})$$

where the values for t_n are listed in Table B.2.

C.2.4 Fading

Prepare and irradiate $i = 1$ to n dosimeters with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to the same value of the conventional true value of the personal dose equivalent, H_t , of at least 1 mSv. Store the dosimeters under standard test conditions for 90 days.

Following storage, etch, read out, determine the measured dose equivalent for each dosimeter, H_i , and calculate the arithmetic mean, $\bar{H}$, and the experimental standard deviation, s . Show that the criterion in Equation (C.11) is met:

$$0,7 \leq (\bar{H} \pm l_i) / H_t \leq 1,3 \quad (C.11)$$

where l_i is calculated in accordance with Equation (B.2).

C.2.5 Ageing

Prepare and store $i = 1$ to n dosimeters under standard test conditions for 90 days. Following storage irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to the same conventional true value of the personal dose equivalent, H_t , of at least 1 mSv.

Etch, read out, determine the measured dose equivalent for each dosimeter, H_i , and calculate the arithmetic mean, $\bar{H}$, and the experimental standard deviation, s . Show that the criterion in Equation (C.12) is met:

$$0,7 \leq (\bar{H} \pm l_i) / H_t \leq 1,3 \quad (C.12)$$

where l_i is calculated in accordance with Equation (B.2).

C.2.6 Exposure to radiation other than neutrons

C.2.6.1 Photon radiation

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters and irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to a conventional true value of the personal dose equivalent, H_t , between 1 mSv and 2 mSv. In addition, irradiate the second lot with 3 mSv of photon radiation from a ^{137}Cs source.

Etch, read out and determine the measured dose equivalent for each dosimeter, H_{ij} , calculate the arithmetic mean, $\bar{H}_j$, and the experimental standard deviation, s_j , for the two lots. Show that the criterion in Equation (C.13) is met:

$$0,9 \leq (\bar{H}_1 / \bar{H}_2) \pm l \leq 1,1 \quad (C.13)$$

where l for the quotient of the arithmetic means is calculated in accordance with Equation (B.5).

C.2.6.2 Radon

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters and expose lot 1 to 3 MBq·h/m³ of radon at 50 % equilibrium with daughters ($F = 0,5$). Store lot 2 in standard test conditions free of radon.

Etch, read out, determine the measured dose equivalent for each dosimeter, H_{ij} , and calculate the arithmetic mean, $\bar{H}_j$, for each of the two lots and the respective experimental standard deviations, s_j . Show that the criterion, expressed in millisieverts, in Equation (C.14) is met:

$$|(\bar{H}_1 - \bar{H}_2)| \pm l \leq 0,5 \quad (C.14)$$

where l for the difference of the arithmetic means is calculated in accordance with Equation (B.5).

C.2.7 Stability under various climatic conditions

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters. Store both lots for 24 h under standard test conditions. Irradiate lot 1 with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to a conventional true value of the personal dose equivalent, H_t , of at least 1 mSv. Store both lots of dosimeters in a climatic chamber in which the temperature is $40\text{ }^{\circ}\text{C} \pm 2\text{ }^{\circ}\text{C}$ and the relative humidity is at least 90 %. After a continuous period of 48 h, remove both lots of dosimeters from the climatic chamber. Irradiate lot 2 to the same conventional true value of the personal dose equivalent as lot 1.

Etch, read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the two lots and the respective experimental standard deviations, s_j . Show that for each lot the criterion in Equation (C.15) is met:

$$0,8 \leq (\bar{H}_j \pm l_{i,j}) / H_t \leq 1,2 \quad (\text{C.15})$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.2.8 Effect of light exposure

C.2.8.1 Effect on zero point

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters and expose lot 1 to $1\,000\text{ W/m}^2$ of light for 1 day. Store dosimeters of lot 2 in the dark in an otherwise identical environment.

Etch, read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the two lots and the respective experimental standard deviations, s_j . Show that the criterion, expressed in millisieverts, in Equation (C.16) is met:

$$|(\bar{H}_1 - \bar{H}_2)| \pm l \leq 1 \quad (\text{C.16})$$

where l for the difference of the arithmetic means is calculated in accordance with Equation (B.5).

C.2.8.2 Effect on the reading

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters and irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source to a conventional true value of the personal dose equivalent, H_t , of at least 1 mSv. Exposure lot 1 to $1\,000\text{ W/m}^2$ of light for 1 week. Store the dosimeters of lot 2 in the dark in an otherwise identical environment.

Etch, read out each dosimeter, $M_{i,j}$, and calculate the arithmetic mean of the reading, $\bar{M}_j$, and the experimental standard deviation, s_j , for the two lots. Show that the criterion in Equation (C.17) is met:

$$0,9 \leq (\bar{M}_1 / \bar{M}_2) \pm l \leq 1,1 \quad (\text{C.17})$$

where l for the quotient of the arithmetic means is calculated in accordance with Equation (B.5).

C.2.9 Energy Dependence of Response

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters and irradiate them with four neutron energies taken from ISO 8529-1:2001, Table 2, in the applicable energy range for which compliance with this standard is sought. The dosimeters shall be irradiated on the ISO slab water phantom to a conventional true value of the personal dose equivalent, H_t , of at least 1 mSv.

Etch, read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (energies) and the respective experimental standard deviation, s_j . Show that the criterion in Equation (C.18) is met:

$$0,5 \leq (\bar{H}_j \pm l_{i,j}) / H_{t,j} \leq 1,5 \quad (C.18)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

NOTE These are the minimum requirements to fully comply with this International Standard. It is possible to use additional reference neutron radiation energies and their use is encouraged.

C.2.10 Angle dependence of response

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters. For this test, a series of mono-energetic neutron energies could be used. It is however convenient that the dosimeters are irradiated on an ISO water slab phantom using a $^{241}\text{Am-Be}$, ^{252}Cf or other neutron source according to the methods of production described in ISO 8529-1. The conventional true value of the personal dose equivalent, H_t , shall be at least 1 mSv and the irradiation conditions shall be the following (ISO 8529-3):

- lot 1: normal incidence;
- lot 2: 15° off normal;
- lot 3: 45° off normal;
- lot 4: 60° off normal.

For lots 2 to 4, the angle of incidence for the n irradiations is varied positive and negative in two planes perpendicular to each other and to the plane of the dosimeter. Determine the conventional true value of the personal dose equivalent by employing the appropriate conversion coefficient for the neutrons used at the respective angles j , $H_{t,j}$.

Etch, read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (angles), the respective experimental standard deviation, s_j , and the response, R_j . Show that the criterion in Equation (C.19) is met:

$$0,6 \leq \left[0,25 \sum_{j=1}^4 (R_j / R_1) \right] \pm l \leq 1,4 \quad (C.19)$$

where l for the sum of the ratios is calculated in accordance with Equation (B.5).

C.3 Thermoluminescence albedo dosimeters (TLAD)

C.3.1 Batch homogeneity

Prepare and identically irradiate each dosimeter from a lot of $i = 1$ to n , dosimeters to a combined neutron/photon field which gives an apparent photon dose equivalent, $H_{ap,i}$, of about 1 mSv to the photon-sensitive detector element and an apparent photon dose equivalent, $H_{an,i}$, of about 2 mSv to the neutron-sensitive element.

Read out and determine the apparent photon dose equivalent, $H_{a,i} = H_{an,i} - H_{ap,i}$, for each dosimeter and show that the coefficient of variation, V , for the n dosimeters does not exceed 20 %.

C.3.2 Reproducibility

Prepare and identically irradiate each dosimeter from a lot of $i = 1$ to n dosimeters to a combined neutron/photon field which gives an apparent photon dose equivalent, $H_{ap,i}$, of about 1 mSv to the photon-sensitive detector element and an apparent photon dose equivalent, $H_{an,i}$, of about 2 mSv to the neutron-sensitive element. Repeat this 10 times with exactly the same value of the apparent photon dose equivalent, H_a , for each of the n dosimeters.

Read out, determine for each of the $i = 1$ to n dosimeters the $j = 1$ to 10 apparent photon dose equivalent, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, the arithmetic mean, $\bar{H}_{a,i}$, and the experimental standard deviation, s_i . Show that for each of the $i = 1$ to n dosimeters, the criterion in Equation (C.20) is met:

$$(s_i + l_{s,i}) / \bar{H}_{a,i} \leq 0,2 \quad (C.20)$$

where the $l_{s,i}$ are calculated in accordance with Equation (B.3).

C.3.3 Linearity

Prepare five lots numbered $j = 1$ to 5 of $i = 1$ to n dosimeters each and irradiate them with a combined neutron/photon field such that the apparent photon dose equivalent to the neutron sensitive element, $H_{an,i}$, is about twice that to the photon-sensitive element, $H_{ap,i}$, and where the conventional true values of the personal dose equivalent, $H_{t,j}$, are 1 mSv, 3 mSv, 10 mSv, 30 mSv and 100 mSv, respectively.

Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each lot j and the respective experimental standard deviations, s_j . Show that for each apparent photon dose equivalent, $H_{a,i,j}$, the criterion in Equation (C.21) is met:

$$0,9 \leq (\bar{H}_j \pm l_{i,j}) / H_{a,i,j} \leq 1,1 \quad (C.21)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.3.4 Detection threshold

Prepare $i = 1$ to n unirradiated dosimeters. Read out, determine the apparent photon dose equivalent, $H_{a,i} = H_{an,i} - H_{ap,i}$, for each dosimeter and calculate the experimental standard deviation, s . Show that the criterion, expressed in millisieverts, in Equation (C.22) is met:

$$t_n \cdot s \leq 0,3^{[27]} \quad (C.22)$$

where the values for t_n are listed in Table B.2.

C.3.5 Fading

Prepare $i = 1$ to n dosimeters and irradiate them all with a combined neutron/photon field such that the apparent photon dose equivalent of the neutron sensitive element, $H_{an,i}$, is about twice that of the photon-sensitive element, $H_{ap,i}$, and where the conventional true value of the personal dose equivalent, H_t , is about 1 mSv. Store the dosimeters under standard test conditions for 90 days. Following storage, read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i} = H_{an,i} - H_{ap,i}$, and calculate the arithmetic mean, $\bar{H}$, and the experimental standard deviation, s . Show that the criterion in Equation (C.23) is met:

$$0,9 \leq (\bar{H} \pm l_i) / H_{a,i} \leq 1,1 \quad (C.23)$$

where l_i is calculated in accordance with Equation (B.2).

C.3.6 Residual signal

Prepare the $i = 1$ to n dosimeters used for the test of the detection threshold (C.3.4) and irradiate them all with a combined neutron/photon field which gives an apparent photon dose equivalent, $H_{ap,i}$, of about 10 mSv to the photon-sensitive detector element and an apparent photon dose equivalent, $H_{an,i}$, of about 20 mSv to the neutron-sensitive element. Anneal and read out the detectors according to the routine procedure.

Anneal and read out the dosimeters a second time, determine the apparent photon dose equivalent for each neutron-sensitive and photon-sensitive detector element, $H_{an,i}$ or $H_{ap,i}$, and calculate the experimental standard deviation, s . Show for each detector element type that the criterion, expressed in millisieverts, in Equation (24) is met:

$$t_n \cdot s \leq 0,3 \quad (C.24)$$

where the values for t_n are listed in Table B.1.

C.3.7 Self-irradiation

Prepare $i = 1$ to n dosimeters. Store them for 60 days under standard test conditions in a location where the background dose rate is known.

Read out, determine the apparent photon dose equivalent for each neutron-sensitive and photon-sensitive detector element, $H_{an,i}$ or $H_{ap,i}$, calculate the arithmetic mean, $\bar{H}_i$, and the experimental standard deviation, s . Show for each detector element type that the criterion, expressed in millisieverts, in Equation (C.25) is met:

$$(\bar{H}_i + l_i) \leq 0,3 \quad (C.25)$$

where l_i is calculated in accordance with Equation (B.2).

C.3.8 Stability under various climatic conditions

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters each. Store both lots for 24 h under standard test conditions. Irradiate lot 1 to a combined neutron/photon field such that the apparent photon dose equivalent of the neutron sensitive element, $H_{an,i,j}$, is about twice that of the photon-sensitive element, $H_{ap,i,j}$. Store both lots of dosimeters in a climatic chamber in which standard test conditions prevail.

After a continuous period of 30 days, remove both lots of dosimeters from the climatic chamber. Irradiate lot 2 in the same way as lot 1. Store both lots for one day under standard test conditions.

Read out, determine the apparent photon dose equivalent, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, for each dosimeter and calculate the arithmetic mean, $\bar{H}_j$, and the experimental standard deviation, s_j , for the two lots. Show that for each lot the criterion in Equation (C.26) is met:

$$0,95 \leq (\bar{H}_j \pm l_{i,j}) / H_{a,i,j} \leq 1,05 \quad (C.26)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

Repeat the test for a storage period of 48 h, but in a climatic chamber in which the temperature is $40^\circ\text{C} \pm 2^\circ\text{C}$ and the relative humidity is at least 90 %. Show that for each lot the criterion in Equation (C.27) is met:

$$0,95 \leq (\bar{H}_j \pm l_{i,j}) / H_{a,i,j} \leq 1,1 \quad (C.27)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.3.9 Effect of light exposure

C.3.9.1 Effect on zero point

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters. Expose lot 1 to 1 000 W/m² of light for one day ensuring that the temperature of the dosimeters is maintained at less than 40 °C. Store the dosimeters of lot 2 in the dark in an otherwise identical environment. Ensure that the temperature of the dosimeters of lot 2 stays within ± 5 °C of the dosimeters of lot 1.

Read out all dosimeters after one day, determine the apparent photon dose equivalent for each neutron-sensitive and photon-sensitive detector element, $H_{an,i,j}$ or $H_{ap,i,j}$, and calculate the arithmetic mean, $\bar{H}_{a,j}$, for each detector element type of the two lots and their respective experimental standard deviations, s_j . Show for each detector element type and for each lot that the criterion, expressed in millisieverts, in Equation (C.28) is met:

$$|\bar{H}_{a,1} - \bar{H}_{a,2}| \pm l \leq 0,3 \quad (\text{C.28})$$

where l is calculated in accordance with Equation (B.5).

C.3.9.2 Effect on the reading

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters each and irradiate them with neutrons to an apparent photon dose equivalent of about 1 mSv. Expose and store lots 1 and 2, respectively, as described in C.3.9.1.

Read out all dosimeters after one week, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each of the two lots and their respective experimental standard deviation, s_j . Show that the criterion in Equation (C.29) is met:

$$0,9 \leq (\bar{H}_1 / \bar{H}_2) \pm l \leq 1,1 \quad (\text{C.29})$$

where l for the ratio of the arithmetic means is calculated in accordance with Equation (B.5).

C.3.10 Angle dependence of response

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters. Irradiate on an ISO water slab phantom using a simulated workplace neutron field in accordance with the methods of production described in ISO 12789 or D₂O-moderated ²⁵²Cf as specified in ISO 8529-1. The conventional true value of the personal dose equivalent, H_t , shall be about 1 mSv and the irradiation conditions shall be the following (ISO 8529-3):

- lot 1: normal incidence;
- lot 2: 15° off normal;
- lot 3: 45° off normal;
- lot 4: 60° off normal.

For lots 2 to 4, the angle of incidence for the n irradiations is varied positive and negative in two planes perpendicular to each other and to the plane of the dosimeter. Determine the conventional true value of the personal dose equivalent employing the appropriate conversion coefficient for the neutrons used in the tests at the respective angles j , $H_{t,j}$.

Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (angles), the respective experimental standard deviation, s_j , and the response, R_j . Show that the criterion in Equation (C.30) is met:

$$0,7 \leq \left[0,25 \sum_{j=1}^4 (R_j / R_1) \right] \pm l \leq 1,3 \quad (C.30)$$

where l for the sum of the ratios is calculated in accordance with Equation (B.5).

C.4 Superheated emulsions (SE)

C.4.1 Batch homogeneity

Ensure that the room temperature is within the specified operating range of the detectors. Prepare and irradiate each of $i = 1$ to n detectors to a specified conventional true value of the personal dose equivalent, H_t , between 0,1 mSv and 1 mSv (preferably to achieve approximately 100 bubbles in the detector) using one of the recommended ISO neutron sources. If the irradiations are done sequentially at a reproducible location, ensure that the room temperature variation during the irradiation period is kept to a minimum (< 2 °C).

Read out each dosimeter in accordance with the manufacturer's recommendation. Evaluate the coefficient of variation, V , for the reading, M_i , for the n dosimeters and show that it does not exceed 25 %.

C.4.2 Reproducibility

Ensure that the room temperature is within the specified operating range of the detectors (20 °C to 40 °C). Prepare and irradiate each of the n detectors of C.4.1 to a specified conventional true value of the personal dose equivalent H_t of 1 mSv or less and repeat this 10 times. The room temperature during irradiation should not vary by more than 2 °C. Read the detectors out in accordance with the manufacturer's recommendation.

For each of the $i = 1$ to n detectors, determine the $j = 1$ to 10 readings, $M_{i,j}$, the arithmetic mean, $\bar{M}_i$, and the experimental standard deviation, s_i . Show that for each of the $i = 1$ to n dosimeters, the criterion in Equation (C.31) is met:

$$(s_i + l_{s,i}) / \bar{M}_i \leq 0,25 \quad (C.31)$$

where the $l_{s,i}$ are calculated in accordance with Equation (B.3).

C.4.3 Linearity

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters each and irradiate them to a conventional true value of personal dose equivalents, $H_{t,j}$, within the range of 0,1 mSv to 1 mSv.

Read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each lot j and the respective experimental standard deviations, s_j . Show that for each personal dose equivalent, $H_{t,j}$, the criterion in Equation (C.32) is met:

$$0,75 \leq (\bar{H}_j \pm l_j) / H_{t,j} \leq 1,25 \quad (C.32)$$


where the l_j are calculated in accordance with Equation (B.2).

C.4.4 Detection threshold

Prepare three lots numbered $j = 1$ to 3 of $i = 1$ to n detectors. Irradiate lot 1 to a conventional true value of the personal dose equivalent, $H_{t,1}$, choosing neutrons in the energy range where the dosimeter is used. The value of $H_{t,1}$ should be approximately 0,05 mSv.

Irradiate lot 2 to approximately $H_{t,2} = 2 H_{t,1}$ and lot 3 to approximately $H_{t,3} = 4 H_{t,1}$. Read the detectors, determine the measured dose equivalent for each dosimeter, $H_{t,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each lot j and the respective experimental standard deviations, s_j .

Fit a straight line through the function $s_j / H_{t,j}$ versus $1/\sqrt{H_{t,j}}$. The required uncertainty in personal dose equivalent $H_{t,j}$ defines a point on the straight-line fit, which gives the value of $1/\sqrt{H_{t,j}}$, from which the required value of $H_{t,j}$ can be determined. This value of $H_{t,j}$ is the detection threshold and shall not exceed 0,1 mSv.



Key

X1 $1/\sqrt{H_{t,j}}$

X2 $s_j / H_{t,j}$

1 relative standard deviation

Figure C.1 — Plot of relative standard deviation on the dose readings, $s_j / H_{t,j}$ versus $1/\sqrt{H_{t,j}}$ for the determination of the detection threshold^[28]

C.4.5 Ageing

C.4.5.1 Superheated drop detectors

Prepare two lots numbered $j = 1, 2$ of $i = 1$ to n superheated drop detectors. Activate the detectors of the first lot and irradiate them under standard conditions to a conventional true value of the personal dose equivalent, H_t , of less than 1 mSv to produce approximately 100 bubbles. Read the detectors, determine the measured dose equivalent for each detector of the first lot, $H_{i,1}$, and calculate the arithmetic mean, $\bar{H}_1$, and the experimental standard deviation, s_1 .

Store the detectors of the second lot in a laboratory away from direct sunlight or other sources of heating, as specified by the manufacturer. Leave the detectors inside their containers for a period of one year.

Repeat the experiment that was carried out on detectors of lot 1 with the detectors of lot 2 under identical conditions. For each of the two irradiations, show that the criterion in Equation (C.33) is met:

$$0,85 \leq (\bar{H}_j \pm l_j) / H_t \leq 1,15 \quad (C.33)$$

where the l_j are calculated in accordance with Equation (B.2).

C.4.5.2 Bubble detectors

Prepare $i = 1$ to n bubble detectors, activate and irradiate the detectors under standard test conditions to a conventional true value of the personal dose equivalent, H_t , of less than 2 mSv to produce approximately 150 bubbles. Following this first irradiation, read the detectors, determine the measured dose equivalent for each detector, $H_{i,1}$, and calculate the arithmetic mean, $\bar{H}_1$, and the experimental standard deviation, s_1 . Recompress the bubbles by screwing the cap back on the detectors 8 h after irradiation. Repeat the experiment j times on a daily basis as per normal work schedule (e.g. 5 days per week).

After continuing the experiment for three months, evaluate the coefficient of variation, V , for the response $\bar{H}_j / H_{t,k}$ of all the measurements j . The coefficient should be less than 25 %.

C.4.6 Residual Signal

Prepare $i = 1$ to n detectors, activate and irradiate the detectors under standard test conditions to a conventional true value of the personal dose equivalent, H_t , of less than 1 mSv to produce approximately 100 bubbles. Reset the detectors in accordance with the manufacturer's recommendation; no bubbles should be present afterwards. Ensure that any new bubbles that may form are not the result of response to natural (i.e. cosmic rays) or radioactive neutron sources.

C.4.7 Stability under various climatic conditions

Prepare $i = 1$ to n detectors. The irradiations are to be performed in controlled temperature environments over the range 20 °C to 40 °C and allow the detectors to reach temperature equilibrium by leaving them in the irradiation area for at least 1 hour. Irradiate the activated detectors to a conventional true value of the personal dose equivalent, H_t , producing approximately 100 bubbles in each detector.

Read out the detectors in accordance with the manufacturer's recommendation, evaluate the measured dose equivalent for each detector, H_i , and determine the arithmetic mean, $\bar{H}$, and the experimental standard deviation, s . Repeat the experiment j times at different temperatures over the range 20 °C to 40 °C. Show that the criterion in Equation (C.34) is met:

$$0,75 \leq (\bar{H}_j \pm l_j) / H_{t,j} \leq 1,25 \quad (C.34)$$

where the l_j are calculated in accordance with Equation (B.2).

C.4.8 Physical damage

Select a lot of $i = 1$ to n detectors. Activate them and then drop them on a representative floor of the working area from a height of 1 m. No more than four bubbles per detector should form.

C.4.9 Energy dependence of response

Prepare m lots numbered $j = 1$ to m of $i = 1$ to n dosimeters and irradiate them with m neutron energies in the applicable energy range for which compliance with this standard is sought. These neutron energies should be taken from ISO 8529-1:2001, Table 2.

Ensure that the irradiations are done under controlled temperature conditions ($< 2^\circ\text{C}$ variation). Activate the detectors and irradiate them on the ISO slab water phantom to a conventional true value of the personal dose equivalent, $H_{t,j}$, producing approximately 100 bubbles in each detector.

Read out the detectors in accordance with the manufacturer's recommendation, evaluate the measured dose equivalent for each detector, $H_{i,j}$, and calculate for each lot j (energy) the arithmetic mean, $\bar{H}_j$, and the experimental standard deviation, s_j . Show that the criterion in Equation (C.35) is met:

$$0,75 \leq (\bar{H}_j \pm l_j) / H_{t,j} \leq 1,25 \quad (\text{C.35})$$

The l_j is calculated in accordance with Equation (B.2).

C.4.10 Angle dependence of response

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n detectors. For this test, a series of mono-energetic neutron energies could be used. It is however convenient that the dosimeters are irradiated on an ISO water slab phantom using $^{241}\text{Am-Be}$, ^{252}Cf or other neutron source according to the methods of production described in ISO 8529-1. The conventional true value of the personal dose equivalent, H_t , shall be at least 1 mSv and the irradiation conditions shall be the following (ISO 8529-3):

- lot 1: normal incidence;
- lot 2: 15° off normal;
- lot 3: 45° off normal;
- lot 4: 60° off normal.

For lots 2 to 4, the angle of incidence for the n irradiations is varied positive and negative in two planes perpendicular to each other and to the plane of the dosimeter. Determine the conventional true value of the personal dose equivalent employing the appropriate conversion coefficient for the neutrons used in the test at the respective angles j , $H_{t,j}$.

Read out, determine the measured dose equivalent for each dosimeter, $H_{i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (angles), the respective experimental standard deviation, s_j , and the response, R_j . Show that the criterion in Equation (C.36) is met:

$$0,7 \leq \left[0,25 \sum_{j=1}^4 (R_j / R_1) \right] \pm l \leq 1,3 \quad (\text{C.36})$$

where l for the sum of the ratios is calculated in accordance with Equation (B.5).

C.5 Ion chamber detectors with direct ion storage

C.5.1 Batch homogeneity

Prepare and identically irradiate each dosimeter from a lot of $i = 1$ to n dosimeters to a combined neutron/photon field which gives an apparent photon dose equivalent, $H_{ap,i}$, of about 1 mSv to the photon-sensitive detector element and an apparent photon dose equivalent, $H_{an,i}$, of about 2 mSv to the neutron-sensitive element.

Read out and determine the apparent photon dose equivalent, $H_{a,i} = H_{an,i} - H_{ap,i}$, for each dosimeter and show that the coefficient of variation, V , for the n dosimeters does not exceed 20 %.

C.5.2 Reproducibility

Prepare and identically irradiate each dosimeter from a lot of $i = 1$ to n dosimeters to a combined neutron/photon field which gives an apparent photon dose equivalent, $H_{ap,i}$, of about 1 mSv to the photon-sensitive detector element and an apparent photon dose equivalent, $H_{an,i}$, of about 2 mSv to the neutron-sensitive element. Repeat this 10 times with exactly the same value of the apparent photon dose equivalent, H_{ap} , for each of the n dosimeters.

Read out, determine for each of the $i = 1$ to n dosimeters the $j = 1$ to 10 apparent photon dose equivalent, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, the arithmetic mean, $\bar{H}_{a,i}$, and the experimental standard deviation, s_i . Show that for each of the $i = 1$ to n dosimeters the criterion in Equation (C.37) is met:

$$(s_i + l_{s,i}) / \bar{H}_{a,i} \leq 0,2 \quad (C.37)$$

where the $l_{s,i}$ is calculated in accordance with Equation (B.3).

C.5.3 Linearity

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters each and irradiate them with a combined neutron/photon field such that the apparent photon dose equivalent to the neutron sensitive element, $H_{an,i,j}$, is about twice that to the photon-sensitive element, $H_{ap,i,j}$, and where the conventional true values of the personal dose equivalent, $H_{t,j}$, are 1 mSv, 2 mSv, 5 mSv and 10 mSv, respectively.

Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, calculate the arithmetic mean, $\bar{H}_j$, for each lot j and the respective experimental standard deviations, s_j . Show that for each apparent photon dose equivalent, $H_{a,i,j}$, the criterion in Equation (C.38) is met:

$$0,9 \leq (\bar{H}_j + l_{i,j}) / H_{a,i,j} \leq 1,1 \quad (C.38)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.5.4 Detection threshold

Prepare $i = 1$ to n unirradiated dosimeters. Read out, determine the apparent photon dose equivalent, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, for each dosimeter and calculate the experimental standard deviation, s . Show that the criterion, expressed in millisieverts, in Equation (C.39) is met:

$$t_n \cdot s \leq 0,3^{[27]} \quad (C.39)$$

where the values for t_n are listed in Table B.2.

C.5.5 Fading

Prepare $i = 1$ to n dosimeters and irradiate them all with a combined neutron/photon field such that the apparent photon dose equivalent of the neutron sensitive element, $H_{an,i}$, is about twice that of the photon-sensitive element, $H_{ap,i}$, and where the conventional true value of the personal dose equivalent, H_t , is about 1 mSv. Store the dosimeters under standard test conditions for 90 days. Following storage, read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i} = H_{an,i} - H_{ap,i}$, and calculate the arithmetic mean, $\bar{H}$, and the experimental standard deviation, s . Show that the criterion in Equation (C.40) is met:

$$0,9 \leq (\bar{H} \pm l_j) / H_{a,i} \leq 1,1 \quad (C.40)$$

where l_j is calculated in accordance with Equation (B.2).

C.5.6 Ageing

Prepare and store $i = 1$ to n dosimeters under standard test conditions for 90 days. Following storage, irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source such that the apparent photon dose equivalent of the neutron sensitive element, $H_{an,i}$, is at least 1 mSv. Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i} = H_{an,i} - H_{ap,i}$, and calculate the arithmetic mean, $\bar{H}$, and the experimental standard deviation, s . Show that the criterion in Equation (C.41) is met:

$$0,9 \leq (\bar{H} \pm l_j) / H_{a,i} \leq 1,1 \quad (C.41)$$

where l_j is calculated in accordance with Equation (B.2).

C.5.7 Exposure to radiation other than neutrons

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters and irradiate them with a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source such that the apparent photon dose equivalent of the neutron sensitive element, $H_{an,i}$, is between 1 mSv and 2 mSv. In addition, irradiate the second lot with 3 mSv of photon radiation from a ^{137}Cs source.

Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, and calculate the arithmetic mean of the reading, $\bar{H}_j$, and the experimental standard deviation, s_j , for the two lots. Show that the criterion in Equation (C.42) is met:

$$0,8 \leq (\bar{H}_1 / \bar{H}_2) \pm l \leq 1,2 \quad (C.42)$$

where l for the quotient of the arithmetic means is calculated in accordance with Equation (B.5).

C.5.8 Stability under various climatic conditions

Prepare two lots numbered $j = 1$ and 2 of $i = 1$ to n dosimeters each. Store both lots for 24 h under standard test conditions. Irradiate lot 1 to a combined neutron/photon field such that the apparent photon dose equivalent of the neutron sensitive element, $H_{an,i}$, is about twice that of the photon-sensitive element, $H_{ap,i}$. Store both lots of dosimeters in a climatic chamber in which the temperature is $40^\circ\text{C} \pm 2^\circ\text{C}$ and the relative humidity is at least 90 %.

After a continuous period of 48 h, remove both lots of dosimeters from the climatic chamber. After waiting for 2 h, irradiate lot 2 in the same way as lot 1.

Read out, determine the apparent photon dose equivalent, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, for each dosimeter and calculate the arithmetic mean, $\bar{H}_j$, and the experimental standard deviation, s_j , for the two lots. Show that for each lot the criterion in Equation (C.43) is met:

$$0,9 \leq (\bar{H}_j \pm l_{i,j}) / H_{a,i,j} \leq 1,1 \quad (C.43)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

C.5.9 Energy dependence of response

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters and irradiate them with four neutron energies taken from of ISO 8529-1:2001, Table 2, in the applicable energy range for which compliance with this standard is sought. The dosimeters shall be irradiated on the ISO slab water phantom to a conventional true value of the personal dose equivalent, H_t , of at least 1 mSv, such that the apparent photon dose equivalent of the neutron detector element is at least twice that of the photon detector element.

Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j} = H_{an,i,j} - H_{ap,i,j}$, apply the relevant calibration procedure to calculate the neutron personal dose equivalent, H_j , and calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (energies) and the respective experimental standard deviation, s_j . Show that the criterion in Equation (C.44) is met:

$$0,5 \leq (\bar{H}_j \pm l_{i,j}) / H_t \leq 1,5 \quad (C.44)$$

where the $l_{i,j}$ are calculated in accordance with Equation (B.2).

NOTE These are the minimum requirements to fully comply with the standard. It is possible to use additional reference neutron radiation energies and their use is encouraged.

C.5.10 Angle dependence of response

Prepare four lots numbered $j = 1$ to 4 of $i = 1$ to n dosimeters. For this test, a series of mono-energetic neutron energies could be used. It is however convenient that the dosimeters are irradiated on an ISO water slab phantom using a $^{241}\text{Am-Be}$ or ^{252}Cf neutron source or other neutron sources according to the methods of production described in ISO 8529-1. The conventional true value of the personal dose equivalent, H_t , shall be at least 1 mSv and the irradiation conditions shall be the following (ISO 8529-3):

- lot 1: normal incidence;
- lot 2: 15° off normal;
- lot 3: 45° off normal;
- lot 4: 60° off normal.

For lots 2 to 4, the angle of incidence for the n irradiations is varied positive and negative in two planes perpendicular to each other and to the plane of the dosimeter. Determine the conventional true value of the personal dose equivalent employing the appropriate conversion coefficient for the neutrons used at the respective angles j , $H_{t,j}$.

Read out, determine the apparent photon dose equivalent for each dosimeter, $H_{a,i,j}$, and calculate the arithmetic mean, $\bar{H}_j$, for each of the four lots (angles), the respective experimental standard deviation, s_j , and the response, R_j . Show that the criterion in Equation (C.45) is met:

$$0,7 \leq \left[0,25 \sum_{j=1}^4 (R_j / R_1) \right] \pm l \leq 1,3 \quad (C.45)$$

where l for the sum of the ratios is calculated in accordance with Equation (B.5).

Annex D (informative)

Irradiation conditions

Reference neutron sources, as specified in ISO 8529-1, that can be traced back to appropriate primary and secondary standards shall be used for type and quality testing of detectors and dosimeters. Radionuclide neutron sources are specified in ISO 8529-1:2001, Table 1, whilst nearly mono-energetic neutron radiations primarily used for determining or testing the energy response of neutron detectors and dosimeters are given in ISO 8529-1:2001, Table 2. These “mono-energetic” neutron sources are available only at highly specialized and well-equipped laboratories staffed with experts in radiation spectroscopy. Generally, these laboratories are already certified by national bodies or are recognized as having the required expertise to investigate the properties of neutron detectors and dosimeters.

Irradiation conditions that shall be observed for the calibration of dosimeters are described in ISO 8529-2 and should also be followed in the type and quality testing of neutron detectors and dosimeters. In particular, the calibration facility should meet the following specifications.

- The total neutron scatter from the floor, walls and ceiling of the room should be kept such that its contribution to the dose equivalent intended to be delivered is smaller than 10 %. Therefore, table and supports should have a minimum mass and not be made of material containing hydrogen. Corrections for scattered neutrons can be made using one of several techniques described in ISO 8529-2.
- If several dosimeters are irradiated together on the same phantom, the distance between them should be sufficiently large that the mutual influence on their reading should be such that the difference in the readings of a dosimeter irradiated together with other dosimeters and the same dosimeter irradiated alone in the same position is less than 3 %.
- In order to subject several dosimeters to the same value of the calibration quantity, their supports should be placed on the same dose rate contour. If sufficient homogeneity cannot be achieved, the support may be made to rotate around the source.

When the dosimeter is calibrated or tested on the ISO water slab phantom, it should be positioned with its back face in contact with the phantom and the radiation impinging perpendicularly onto the surface of the dosimeter, except in the case where the angle dependence of the response is determined. The “source to dosimeter” distance is defined as the distance from the equivalent point source to the geometric centre of the sensitive volume of the dosimeter. The necessary neutron fluence-to-dose equivalent conversion coefficients $h_{p\phi}(10;E,\alpha)$ are given in ISO 8529-3 for both radionuclide neutron sources and nearly mono-energetic neutron radiations as specified in ISO 8529-1.

Simulated workplace neutron fields as specified and produced in accordance with the methods described in ISO 12789 can also be chosen for testing and calibration, provided these spectra are sufficiently characterized in term of their spectral fluence. This will permit the establishment of a fluence-to-dose equivalent conversion coefficient for the spectrum of interest and define reference dose equivalents at the calibration position. Simulated workplace neutron fields are important in practical radiation protection because they closely resemble neutron fields encountered in routine radiation protection. However, for type tests of angle dependence, simulated workplace neutron fields are only recommended for thermoluminescence albedo dosimeters. Complete information on these fields and the “field” calibration conditions shall be contained in adequate certification documents.

Annex E (informative)

Performance of neutron dosimeters

All tests are performed on a specified number of dosimeters randomly selected. The detectors shall be chosen from a single production batch, i.e. from one emulsion number in the case of the nuclear track emulsion or material from the same sheet of material in the case of the solid-state nuclear track detectors.

The number of dosimeters used in each test shall be such to demonstrate that the performance requirements are met at the 95 % confidence level. The performance tests proposed take in consideration the particular features of the detectors used. Hence, in the case of the nuclear emulsion and the solid state nuclear track detector, a test on reproducibility is meaningless, as the detectors cannot be reused. On the other hand, the following tests are pertinent to all dosimeters.

Batch homogeneity is determined by observing the variability among the readings of dosimeters subjected to the same value of the calibration quantity under identical conditions, including laboratory and operator.

The read out value of a dosimeter should be linearly related to the personal dose equivalent, $H_p(10)$, over its whole useful dose range.

The detection threshold of any detector is a function of the variability of the background that shall be determined for each batch of detectors or dosimeters.

Temperature, humidity and light influence the response of most dosimeters. Tests with respect to temperature and humidity are performed in climatic chambers. To produce 1 000 W/m² of light, use an apparatus which produces light where the spectrum corresponds to that of bright sunlight (295 nm to 769 nm). This is achieved, for example, with a xenon lamp equipped, if necessary, with appropriate filters or by using a daylight fluorescent lamp.

NOTE 1 000 W/m² of bright sunlight includes nominally 60 W/m² of U.V.

In addition to the angle dependence of the response, the variation in response to personal dose equivalent, $H_p(10)$, with neutron energy is one of the most important performance parameters of the dosimeters to be tested as all known personal neutron dosimeters show a strong dependence of response with neutron energy.

E.1 Nuclear track emulsions

The nuclear track emulsion (NTE), the oldest passive method for neutron detection, is used in personal dosimetry for persons working with radioactive neutron sources and in stray neutron fields outside the biological shielding around accelerators where neutron energies reach up to the maximum energy of the primary beam. The nuclear emulsion is no longer employed around nuclear power reactors where neutron energies are mostly below the detection threshold of the detector.

The loss of latent information or fading with time is the greatest problem in nuclear track emulsions. In order to reduce the fading to a minimum, the nuclear emulsion is, before use, generally desiccated in dry nitrogen and sealed in an air- and moisture-proof pouch called a film packet. The test on fading checks the stability of the latent image under normal test conditions and should assure that the dosimetric information stored stays sufficiently stable during a normal wearing period.

The packaging and the development of the nuclear track emulsion as well as the scanning equipment used influence the performance.

Since fading of nuclear tracks is strongly influenced by temperature and humidity, the quality of the air- and water-tight pouch used to seal the nuclear track emulsion shall be checked at higher temperature and humidity, i.e. conditions that correspond to the most adverse practical wearing practices of the dosimeter.


Nuclear track emulsions are carried as personal dosimeters in radiation fields containing photons. A fogging of the emulsion by an important exposure to photons will make the scanning of recorded neutron tracks difficult. For a reasonable exposure to photons, generally in mixed neutron-gamma radiation fields, the loss of information for the neutron component should be negligible.

The film package shall be light-tight and shall be tested accordingly.

Due to the fact that in the case of the nuclear track emulsion at least 100 tracks per dosimeter must be counted to reach a sufficiently low stochastic uncertainty, it is inconvenient to use too many dosimeters. However, all tests should be performed with lots comprising at least five dosimeters.

The tests on the angle and the energy dependence of the response shall prove the compliance of the nuclear track emulsion as a dosimeter measuring the quantity personal dose equivalent, $H_p(10)$, with respect to angle and energy of a neutron exposure with sufficient accuracy.

For nuclear track emulsions, the useful neutron energy range starts between 500 keV and 1 MeV with its response to personal dose equivalent rising to a maximum at 14 MeV, as shown in Figure E.1.



NOTE Open symbols are taken from Reference [4], whilst solid symbols are from Reference [13]. It is evident that for tests, only those neutron sources in the energy range from 2 MeV to 5 MeV and chosen from ISO 8529-1 are meaningful and should be used. However, for convenience, all tests could be made with neutrons from readily available $^{241}\text{Am-Be}$ or ^{252}Cf sources, except the one on energy response.

Figure E.1 — Energy response for the nuclear track emulsion in tracks per mm^2 and per mSv