
**Horizontal methods for molecular
biomarker analysis — Methods
of analysis for the detection of
genetically modified organisms and
derived products —**

**Part 3:
Construct-specific real-time PCR
method for detection of P35S-pat-
sequence for screening genetically
modified organisms**

*Méthodes horizontales d'analyse moléculaire de biomarqueurs —
Méthodes d'analyse pour la détection des organismes génétiquement
modifiés et des produits dérivés —*

*Partie 3: Méthode PCR en temps réel spécifique de la construction
pour la détection de la séquence P35S-pat pour criblage des
organismes génétiquement modifiés*



STANDARDSISO.COM : Click to view the full PDF of ISO/TS 21569-3:2015



COPYRIGHT PROTECTED DOCUMENT

© ISO 2015

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO at the address below or ISO's member body in the country of the requester.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

Published in Switzerland

Contents

| | Page |
|---|----------|
| Foreword | iv |
| 1 Scope | 1 |
| 2 Normative references | 1 |
| 3 Terms and definitions | 1 |
| 4 Principle | 1 |
| 5 Reagents and materials | 2 |
| 5.1 General | 2 |
| 5.2 PCR reagents | 2 |
| 6 Apparatus | 3 |
| 7 Procedure | 3 |
| 7.1 Test sample preparation | 3 |
| 7.2 Preparation of the DNA extracts | 3 |
| 7.3 PCR setup | 3 |
| 7.4 Temperature-time programme | 4 |
| 8 Accept/reject criteria | 4 |
| 8.1 General | 4 |
| 8.2 Identification | 4 |
| 8.3 Calculation of <i>P35S-pat</i> copy numbers | 4 |
| 9 Validation status and performance criteria | 5 |
| 9.1 Robustness of the method | 5 |
| 9.2 Collaborative trial for determination of LOD | 5 |
| 9.3 Collaborative trial for quantification of the <i>P35S-pat</i> construct in rapeseed | 6 |
| 9.4 Sensitivity | 7 |
| 9.5 Specificity | 7 |
| 10 Test report | 8 |
| Bibliography | 9 |

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 34, *Food products*, Subcommittee SC 16, *Horizontal methods for molecular biomarker analysis*.

ISO 21569 consists of the following parts, under the general title *Horizontal methods for molecular biomarker analysis — Methods of analysis for the detection of genetically modified organisms and derived products*:

- *Part 2: Construct-specific real-time PCR method for detection of event FP967 in linseed and linseed products* [Technical Specification]
- *Part 3: Construct-specific real-time PCR method for the detection of the P35S-pat-sequence for screening for compounds of genetically modified organisms* [Technical Specification]

ISO 21569:2005 is to be revised to become the future Part 1.

Horizontal methods for molecular biomarker analysis — Methods of analysis for the detection of genetically modified organisms and derived products —

Part 3:

Construct-specific real-time PCR method for detection of P35S-pat-sequence for screening genetically modified organisms

1 Scope

This Technical Specification describes a procedure for the detection of the DNA transition sequence between the 35S promoter (*P35S*) from *Cauliflower mosaic virus* and a modified phosphinothricin-acetyltransferase gene (*pat*) from *Streptomyces viridochromogenes*. The *P35S-pat* construct is frequently found in genetically modified plants with tolerance for phosphinothricin-containing herbicides. The *P35S-pat* construct specific method is based on a real-time PCR and can be used for qualitative and quantitative screening purposes. For identification and quantification of a specific event, a follow-up analysis has to be carried out.

This Technical Specification is applicable for the analysis of DNA extracted from foodstuffs. It may also be suitable for the analysis of DNA extracted from other products such as feedstuffs and seeds. The application of this method requires the extraction of an adequate quantity and quality of amplifiable DNA from the relevant matrix.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 21569, *Foodstuffs — Methods of analysis for the detection of genetically modified organisms and derived products — Qualitative nucleic acid based methods*

ISO 21571, *Foodstuffs — Methods of analysis for the detection of genetically modified organisms and derived products — Nucleic acid extraction*

ISO 24276, *Foodstuffs — Methods of analysis for the detection of genetically modified organisms and derived products — General requirements and definitions*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 24276 apply.

4 Principle

DNA is extracted from the test portion applying a suitable method. The DNA analysis consists of two parts, namely,

- 1) verification of the amount and amplifiability of the extracted DNA, e.g. by means of a target taxon specific real-time PCR (see ISO 21570^[10]), and

2) detection of the *P35S-pat* construct in a real-time PCR.^[1,2]

5 Reagents and materials

5.1 General

Chemicals of recognized analytical grade, appropriate for molecular biology shall be used, as a rule. The water used shall be double distilled or PCR grade water (i.e. nuclease and nucleic acid free). For all operations in which gloves are used, it should be ensured that these are powder-free. The use of aerosol-protected pipette tips as protection against cross contamination is recommended.

5.2 PCR reagents

5.2.1 Thermostable DNA polymerase, for hot-start PCR.

5.2.2 PCR buffer solution, containing magnesium chloride and deoxyribonucleoside triphosphates (dATP, dCTP, dGTP and dUTP).

Ready-to-use reagent mixtures or mixes of individual components can be used. Reagents and polymerases which lead to equal or better results may also be used.

5.2.3 Oligonucleotides (see [Table 1](#)).

Table 1 — Oligonucleotides

| Name | DNA sequence of the oligonucleotide | Final concentration in the PCR |
|--|---|--------------------------------|
| <i>P35S-pat</i> construct as the target sequence: ^[1,2] | | |
| Primer 35SP03.f | 5'-AAG TTC ATT TCA TTT ggA gAg gAC A-3' | 200 nmol/l |
| Primer pat-7.r | 5'-Cgg CCA TAT CAg CTg CTg TAG-3' | 200 nmol/l |
| Probe GSS01.s | 5'-(FAM)-CCg gAg Agg AgA CCA gTT gAg ATT Agg C-(TAMRA)-3'a | 100 nmol/l |
| ^a FAM: 6-Carboxyfluorescein, TAMRA: 6-Carboxytetramethylrhodamine | | |

NOTE Equivalent reporter dyes and/or quencher dyes can be used for the probe if they can be shown to yield similar or better results.

5.2.4 Standard DNA for calibration

A standard DNA solution of a known concentration (ng/μl) can be used to calculate the copy number of the *P35S-pat* target sequence.

When using genomic plant DNA as the standard DNA, the number of haploid genome equivalents should be calculated on the basis of the molecular mass of the plant haploid genome by applying the Formula (1):

$$\text{Number of genome equivalents per } \mu\text{l} = \frac{\text{DNA concentration [ng/}\mu\text{l}] \cdot 1\,000}{\text{haploid genome mass [pg]}} \quad (1)$$

On the basis of the genome equivalents, the respective copy number for the *P35S-pat* sequence can be calculated. In doing so, the number of integrations into the plant genome as well as the degree of zygosity of the plant material used shall be taken into consideration.

6 Apparatus

Regarding the apparatus and materials, see ISO 21569. In addition to the usual laboratory equipment, the following equipment is required.

6.1 Real-time PCR device, suitable for the excitation of fluorescent molecules and the detection of fluorescence signals generated during PCR.

7 Procedure

7.1 Test sample preparation

It should be ensured that the test sample used for DNA extraction is representative of the laboratory sample, e.g. by grinding or homogenizing of the samples. Measures and operational steps to be taken into consideration shall be as described in ISO 21571 and ISO 24276.

7.2 Preparation of the DNA extracts

Concerning the preparation of DNA from the test portion, the general instructions and measures described in ISO 21571 should be followed. It is recommended to choose one of the DNA extraction methods described in ISO 21571, Annex A.

7.3 PCR setup

The method is described for a total volume of 25 µl per PCR. The reaction set-up is given in [Table 2](#).

Completely thaw reagents at room temperature. Each reagent should be carefully mixed and briefly centrifuged immediately before pipetting. Prepare a PCR reagent mixture which contains all components except for the sample DNA. The required amount of the PCR reagent mixture depends on the number of reactions to be performed, including at least one additional reaction as a pipetting reserve. Add 5 µl of sample DNA to each reaction.

Table 2 — Reaction set-up for the amplification

| | |
|---|-----------------------------|
| Overall reaction volume | 25 µl |
| Sample DNA (up to 200 ng) or controls | 5 µl |
| PCR buffer solution ^a (including MgCl ₂ , dNTPs and hot-start DNA polymerase) | 12,5 µl |
| Primer 35SP03.f and pat-7.r | see Table 1 |
| Probe GSS01.s | see Table 1 |
| Water | To 25 µl |
| ^a In the collaborative study, depending on the real-time PCR devices, different PCR buffer solutions were used [TaqMan Universal PCR Mastermix (Life Technologies, Darmstadt), QuantiTect Multiplex PCR NoROX or QuantiTect Probe PCR Mastermix (Qiagen GmbH, Hilden)]. This information is given for the convenience of users of this document and does not constitute an endorsement by ISO. Equivalent products from other manufacturers may be used if they yield similar or better results. If necessary, adapt the amounts of the reagents and the temperature-time programme. | |

Mix the PCR reagent mixture, centrifuge briefly and pipette 20 µl into each reaction vial. For the amplification reagent control, add 5 µl water into the respective reaction set-up. Pipette either 5 µl of sample DNA or 5 µl of the respective control solution (extraction blank control, positive DNA target control). If necessary, prepare a PCR inhibition control as described in ISO 24276.

Transfer the reaction set-ups into the thermal cycler and start the temperature-time programme.

7.4 Temperature-time programme

The temperature-time programme as outlined in [Table 3](#) has been used in the validation study. The use of different reaction conditions and real-time PCR cyclers may require specific optimization. The time for initial denaturation depends on the master mix used.

Table 3 — Temperature-time programme

| Step | Parameter | Temperature | Time | Fluorescence measurement | Cycles |
|------|---------------------------|--------------------------|--------|--------------------------|--------|
| 1 | UNG activation (optional) | 50 °C | 2 min | no | 1 |
| 2 | Initial denaturation | 95 °C | 10 min | no | 1 |
| 3 | Amplification | Denaturation | 95 °C | no | 45 |
| | | Annealing and elongation | 60 °C | yes | |

8 Accept/reject criteria

8.1 General

A corresponding real-time PCR device-specific data analysis programme is used for the identification of PCR products. The amplification results may be expressed in a different manner, depending on the device used. In the absence of detectable PCR products (e.g. negative control), the result can be expressed as 'undetermined', 'no amp', or the maximum number of possible cycles. If the amplification of the DNA target sequence occurs in a sample (e.g. positive control), a sigmoid-shaped amplification curve can be observed and the cycle number is calculated at which a predetermined fluorescence threshold value is exceeded (C_t value or C_p value).

If, due to atypical fluorescence measurement data, the automatic interpretation does not provide a meaningful result, it might be necessary to set the baseline and the threshold manually prior to interpreting the data. In this case, the device-specific instructions given in the manual regarding the use of the interpretation software shall be applied.

8.2 Identification

The target sequence is considered as detected, if

- by using the *P35S-pat* specific primers 35SP03.f and pat-7.r and the probe GSS01.s, a sigmoid-shaped amplification curve can be observed and a predetermined fluorescence threshold value is exceeded,
- in the PCR control set-ups with no added DNA (PCR reagent control, negative extraction control), no sigmoid shaped amplification curve can be observed and a predetermined fluorescence threshold value is not exceeded, and
- in the set-ups for the amplification control (positive DNA target control, PCR inhibition control) the expected C_t values (or C_p values) are achieved.

8.3 Calculation of *P35S-pat* copy numbers

With DNA standards containing defined amounts of *P35S-pat* copy numbers (see [5.2.4](#), [9.3](#)), a calibration curve can be generated and used for calculation of *P35S-pat* copy numbers in unknown samples.

9 Validation status and performance criteria

9.1 Robustness of the method

The robustness of the method has not been tested with respect to small modifications of factors such as reagent concentrations (e.g. primers, probe) or reaction conditions (e.g. annealing temperature).

NOTE In the collaborative trial the robustness of the method has been checked with regard to different real-time PCR machines and PCR buffer solutions. The real-time PCR machines and the PCR buffer solutions had no influence on the performance of the method.

9.2 Collaborative trial for determination of LOD

The LOD of the method has been tested in a collaborative study coordinated in 2011 by the German Federal Office of Consumer Protection and Food Safety (BVL) with a total of 10 participants. The participants received 4 DNA samples from GM-positive plants containing the *P35S-pat* target sequence.

To prepare the samples, certified reference materials produced from the American Oil Chemists' Society (AOCS, Urbana USA) and the Institute for Reference Materials and Measurements (IRMM, Geel Belgium) were used. The materials used were DNA from leaves of T45 canola (AOCS, 0208-A2), of A2704-12 soybeans (AOCS, 0707-B2), of T25 maize (AOCS, 0306-H), and DNA extracted from 1507 maize powder (IRMM, ERM-BF418d). DNA concentrations were determined photospectrometrically. Numbers of genome equivalents per μl were calculated applying Formula (1). On the basis of the number of genome equivalents, the respective copy number for the *P35S-pat* sequence was calculated considering the number of integrations of the *P35S-pat* sequence into the plant genome as well as the degree of zygosity of the plant material used (see Table 4). Copy numbers of sample DNAs were adjusted to approximately 100 copies of the *P35S-pat* sequence per μl .

Table 4 — Characteristics of reference materials for determination of LOD

| GM event | Source of reference material | Material/ % (m/m) GM | Haploid genome mass (pg) [3] | Zygosity | <i>P35S-pat</i> copy numbers per haploid genome | Length of the <i>P35S-pat</i> PCR product (bp) |
|------------------|------------------------------|-------------------------|---------------------------------|-------------|---|--|
| T45 canola | AOCS | DNA / 99,99 % | 1,3 | homozygot | 1[4] | 111 |
| A2704-12 soybean | AOCS | DNA / 99,99 % | 1,13 | homozygot | 2[5] | 102 |
| T25 maize | AOCS | DNA / 99,99 % | 2,7 | homozygot | 1[6] | 111 |
| TC1507 maize | IRMM | powder / 10 % | 2,7 | heterozygot | 1[7] | 102 |

On the basis of these four standard DNA solutions, dilution series were prepared by the participants in order to obtain DNA solutions with copy numbers of 50 to 0,1 per 25 μl PCR reaction, respectively. The participants analysed each diluted DNA solution with the *P35S-pat* real-time PCR method in a sixfold determination under the conditions described in Table 1 to Table 3. The results of this collaborative trial study are listed in Table 5.

Table 5 — Results of the collaborative trial for determination of LOD

| <i>P35S-pat</i> copy numbers per PCR | Number of positive results ($C_t < 45$) out of 60 results | | | |
|--------------------------------------|---|--------------|-----------|------------------|
| | T45 canola | TC1507 maize | T25 maize | A2704-12 soybean |
| 50 | 60 | 60 | 60 | 60 |
| 20 | 60 | 60 | 60 | 60 |
| 10 | 59 | 60 | 60 | 60 |
| 5 | 46 | 59 | 56 | 55 |

Table 5 (continued)

| <i>P35S-pat</i> copy numbers per PCR | Number of positive results ($C_t < 45$) out of 60 results | | | |
|--------------------------------------|---|--------------|-----------|------------------|
| | T45 canola | TC1507 maize | T25 maize | A2704-12 soybean |
| 2 | 26 | 52 | 41 | 36 |
| 1 | 11 | 42 | 21 | 24 |
| 0,1 | 3 | 8 | 3 | 2 |

9.3 Collaborative trial for quantification of the *P35S-pat* construct in rapeseed

The method has been validated in a collaborative study coordinated in 2004 by the Sub-committee for Method Development of the German National and Federal Joint Committee on Genetic Engineering (LAG) with a total of 14 participants. The participants received 5 seed samples and 5 DNA samples, respectively.

For preparation of the seed samples, non-GM rapeseed was mixed with genetically modified GS40/90 rapeseed in ratios of 980 g / 20 g (2 % GM), 990 g / 10 g (1 % GM), 995 g / 5 g (0,5 % GM), 999 g / 1 g (0,1 % GM), and 1 000 g / 0 g (0 % GM), homogenized and subdivided into samples of 30 g using a sample divider. Participants had to grind the seed samples and isolate the DNA thereof. For preparation of the DNA samples, provided to the participants, genomic DNA isolated from a plant of GS40/90 oilseed rape was mixed with genomic DNA isolated from a plant of non-GM oilseed rape to obtain solutions containing 2 %, 1 %, 0,5 %, 0,1 %, and 0 % of GS40/90 oilseed rape DNA. Concentrations of the DNA samples were adjusted to 20 ng/μl.

Each sample was analysed by the participants in a threefold determination with the *P35S-pat* real-time PCR method under the conditions described in Table 1 to Table 3. For calibration 6 DNA standards (100, 300, 900, 2 700, 8 100, 100 000 copies; provided) were measured twofold in the same PCR analysis run. In addition, the same number of measurements was carried out using a taxon specific real-time PCR method targeting the *pepC* gene of oilseed rape [1]. The calibration curves were created by plotting the C_t values against the logarithm of the copy numbers of the target sequence provided for the calibration solutions. The respective copy numbers for the samples were calculated by interpolation from the calibration curve.

In Table 6, a summary of the results is presented. Table 7 shows the quantitative results. Before the calculation of the mean GM contents and of precision data using statistical tests according to ISO 5725-2 [8], Grubbs and Cochran outliers were identified and eliminated if applicable. Since the number of *pepC* alleles in different varieties of oilseed rape is not known, it is not possible to use the results of the taxon specific PCR to calculate the oilseed rape genome copy number. Therefore, the trueness of the method could not be determined. In addition, it is assumed that the true GM content in the seed samples may deviate from the expected content. The degree of deviation cannot be estimated.

Table 6 — Results of the collaborative trial for quantification of the *P35S-pat* construct in rapeseed

| | |
|--|------|
| Year of collaborative trial | 2004 |
| Number of laboratories | 14 |
| Number of laboratories submitting results | 14 |
| Number of seed samples per laboratory | 5 |
| Number of DNA samples per laboratory | 5 |
| Number of accepted results | 140 |
| Number of accepted samples containing the <i>P35S-pat</i> target sequence | 112 |
| Number of accepted samples which did not contain the <i>P35S-pat</i> target sequence | 28 |

Table 6 (continued)

| | |
|------------------------|---------|
| False-positive results | 0 (0 %) |
| False-negative results | 0 (0 %) |

Table 7 — Quantitative results obtained in the *P35S-pat* collaborative trial

| Relative GM contents of samples in % (w/w) | Calculated GM contents in % (<i>P35S-pat</i> copies / <i>pepC</i> copies) | |
|--|---|-----------------------------------|
| | MW ^a | RSD _R ^b (%) |
| 0,1 % (rapeseed) | 0,19 | 52,0 |
| 0,5 % (rapeseed) | 0,88 | 29,5 |
| 1 % (rapeseed) | 1,78 | 42,1 |
| 2 % (rapeseed) | 3,91 | 32,5 |
| 0,1 % (DNA) | 0,14 | 32,2 |
| 0,5 % (DNA) | 0,54 | 34,5 |
| 1 % (DNA) | 1,08 | 21,8 |
| 2 % (DNA) | 2,26 | 17,3 |
| ^a Mean value after elimination of Grubbs and Cochran outliers | | |
| ^b Relative standard deviation (under reproducibility conditions) after elimination of Grubbs and Cochran outliers | | |

9.4 Sensitivity

In Table 5, the collaborative trial results for DNA samples with low copy numbers of the *P35S-pat* target sequence are shown. The standard DNA, set to a concentration of 10 copies per PCR reaction, resulted in an amplification in almost all laboratories with only one exception out of 60 reactions for T45 canola. Based on these results, the limit of detection for the tested GM events is less than or equal to 10 copies.

9.5 Specificity

The specificity of the primer and the probe was validated *in silico* using sequence alignments of data searches in GenBank/EMBL/DDBJ. For this purpose, by use of the programme BLASTN and the sequence of the PCR product of maize event DAS-59122-7 (GenBank accession FN550388, length 102 bp), a search for matches was performed in both the GenBank nucleotide sequence collection ('non-redundant' database with all GenBank, RefSeq, EMBL, DDBJ, and PDB sequences) and the database for patented nucleotide sequences. The results of the searches show no complete identity with other sequences in the databases except for those targeted by the oligonucleotides. Identity of the amplicon sequence occurs for patented nucleotide sequences of DAS 59122-7, DAS-59132, TC1507, and A5547-127, as well as for other patented nucleotide sequences of not-specified GM plants or vectors used for transformation of plants. A BLASTN search (search date: 15 November 2011) confirmed complete identity of the *P35S-pat* sequence only to nucleotide sequences of GM plants or vectors used for transformation of plants.

When experimentally determining the specificity of the *P35S-pat* method, no amplification was detected with DNA from the following GM plants:

- **GM rice:** LL62 (ACS-OS002-5), LL601 (BCS-OS003-7);
- **GM rapeseed:** GT73 (MON-00073-7), Laurat (pCGN3828) (CGN-89465-2), MS1xRF1 (ACS-BN004-7xACS-BN001-4), MS8 (ACS-BN005-8);
- **GM maize:** 3272 (SYN-E3272-5), Bt176 (SYN-EV176-9), CBH-351 (ACS-ZM004-3), MON809 (PH-MON809-2), MON810 (MON-00810-6), GA21 (MON-00021-9), MIR604 (SYN-IR604-5), MON863 (MON-00863-5), MON88017 (MON-88017-3), NK603 (MON-00603-6);
- **GM soy:** 305423 (DP-305423-1), 356043 (DP-356043-5), MON40-3-2 (MON-04032-6);